



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107253

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/20/2020	12:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/20/2020	12:01
3	CAL3	1	Water	NA	NA	20.1	10.04	01/20/2020	12:02
4	ICV	1	Water	NA	NA	20.1	7.00	01/20/2020	12:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/20/2020	12:06
6	L1187-01	1	Solid	20.03	20	20.9	8.67	01/20/2020	12:08
7	L1187-01DUP	1	Solid	20.04	20	20.8	8.68	01/20/2020	12:09
8	CCV2	1	Water	NA	NA	20.1	12.01	01/20/2020	12:10

28107253

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1187 S WorkList ID : 135187 Department : Wet-Chemistry Date : 01-20-2020 11:48:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/17/2020	Solid	L1187-01 -18	pH	Cool 4 deg C	NEWP02	I42	STOCKPILE	01/17/2020	9045D

Date/Time 01-20-2020 11:50
Received by: JS
Relinquished by: DL

Date/Time 01-20-2020 12:35
Received by: DK
Relinquished by: Ja