



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107254

Slope : 99.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 01/20/2020 | 12:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 01/20/2020 | 12:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.04       | 01/20/2020 | 12:02     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 01/20/2020 | 12:05     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 01/20/2020 | 12:15     |
| 6   | L1202-01    | 1  | Water  | NA          | NA          | 18.8             | 7.41        | 01/20/2020 | 12:16     |
| 7   | L1202-01DUP | 1  | Water  | NA          | NA          | 18.9             | 7.42        | 01/20/2020 | 12:17     |
| 8   | L1202-02    | 1  | Water  | NA          | NA          | 19.3             | 7.51        | 01/20/2020 | 12:20     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.2             | 12.01       | 01/20/2020 | 12:22     |

28107254

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1202      WorkList ID : 135186      Department : Wet-Chemistry      Date : 01-20-2020 11:36:22

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|-------------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 01/20/2020 | Water  | L1202-01 -D | pH   | Cool 4 deg C | TULL02   | I51              | F1-F2-(SETTLED) | 01/20/2020   | 9040C  |
| 01/20/2020 | Water  | L1202-02 -D | pH   | Cool 4 deg C | TULL02   | I51              | E2-E3-(SETTLED) | 01/20/2020   | 9040C  |

Date/Time 01-20-2020 11:40  
Received by: Jg  
Relinquished by: Jg

Date/Time 01-20-2020 12:35  
Received by: Jg  
Relinquished by: Jg