



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107255

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP81353
Turbidity Calibration std, 0NTU	WP81352
Turbidity Calibration std, 20NTU	WP81348
Turbidity Calibration std, 40NTU	WP81347
Turbidity Calibration std, 1NTU	WP81351
Turbidity Calibration std, 10NTU	WP81349
Turbidity Calibration std, 5NTU	WP81350
10 NTU Standard 500 ml	W2489

Intercept: -0.1169

Slope: 0.9183

Regression: 0.9992

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.160	01/20/2020	14:20
2	CAL2	1	1	0.992	01/20/2020	14:21
3	CAL3	5	1	4.380	01/20/2020	14:23
4	CAL4	10	1	9.260	01/20/2020	14:26
5	CAL5	20	1	17.200	01/20/2020	14:28
6	CAL6	40	1	37.100	01/20/2020	14:30



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107255

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.100	10.037	01/20/2020	14:32
2	ICB		1	0.151	0.292	01/20/2020	14:33
3	CCV1	10	1	9.200	10.146	01/20/2020	14:35
4	CCB1		1	0.146	0.286	01/20/2020	14:37
5	LB107255BL		1	0.149	0.290	01/20/2020	14:39
6	L1173-03		1	20.200	22.124	01/20/2020	14:43
7	L1173-03DUP		1	19.600	21.471	01/20/2020	14:45
8	CCV2	10	1	9.190	10.135	01/20/2020	14:47
9	CCB2		1	0.139	0.279	01/20/2020	14:49

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tur01202020 WorkList ID : 135192 Department : Wet-Chemistry Date : 01-20-2020 13:57:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2020	Water	L1173-03	Turbidity	Cool 4 deg C	AQUA04	I31	0052-DECANT-H2O-COMP-4	01/16/2020	SM2130 B

Date/Time 01-20-2020 14:01

Received by: AK

Relinquished by: JP

Date/Time 01-20-2020 14:35

Received by: JP

Relinquished by: AK