

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB107257

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP81300
calibration std. phosphate 0.5 ppm	WP81299
calibration std. phosphate 0.3 ppm	WP81298
calibration std. phosphate 0.1 ppm	WP81297
calibration std. phosphate 0.05 ppm	WP81296
calibration std. 0 ppm	WP81294
phosphate ICV std.	WP81302
phosphate CCV std.	WP81301
5N sulfuric acid	WP79628
Combined reagent	WP81336
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314

Intercept: -0.0007Slope: 0.6545Regression: 0.999868

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/20/2020	13:45
2	CAL2	0.05	1	50	50	0.036	01/20/2020	13:45
3	CAL3	0.10	1	50	50	0.066	01/20/2020	13:46
4	CAL4	0.30	1	50	50	0.188	01/20/2020	13:46
5	CAL5	0.50	1	50	50	0.326	01/20/2020	13:47
6	CAL6	1.00	1	50	50	0.656	01/20/2020	13:47



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB107257

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.306	0.469	01/20/2020	13:48
2	ICB		1	50	50	0.000	0.001	01/20/2020	13:49
3	CCV1	0.50	1	50	50	0.325	0.498	01/20/2020	13:49
4	CCB1		1	50	50	0.000	0.001	01/20/2020	13:50
5	PB126213BL		1	50	50	0.000	0.001	01/20/2020	13:50
6	PB126213BS	0.50	1	50	50	0.310	0.475	01/20/2020	13:51
7	L1183-01		1	50	50	2.140	3.271	01/20/2020	13:51
8	L1183-01DUP		1	50	50	1.971	3.013	01/20/2020	13:52
9	L1183-01MS	0.50	1	50	50	2.216	3.387	01/20/2020	13:52
10	L1183-01MSD	0.50	1	50	50	2.240	3.424	01/20/2020	13:53
11	L1183-01		50	50	50	0.260	0.398	01/20/2020	14:01
12	L1183-01DUP		50	50	50	0.261	0.400	01/20/2020	14:01
13	L1183-01MS	0.50	50	50	50	0.266	0.407	01/20/2020	14:02
14	L1183-01MSD	0.50	50	50	50	0.265	0.406	01/20/2020	14:03
15	CCV2	0.50	1	50	50	0.328	0.502	01/20/2020	14:03
16	CCB2		1	50	50	0.000	0.001	01/20/2020	14:04