



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107266

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/20/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/20/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/20/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	01/20/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/20/2020	17:06
6	L1004-15	1	Solid	20.03	20	21.1	7.66	01/20/2020	17:10
7	L1004-15DUP	1	Solid	20.04	20	21.2	7.67	01/20/2020	17:11
8	L1210-07	1	Solid	20.02	20	21.6	8.65	01/20/2020	17:12
9	L1211-02	1	Solid	20.03	20	21.3	7.51	01/20/2020	17:15
10	L1211-04	1	Solid	20.05	20	21.2	8.48	01/20/2020	17:16
11	CCV2	1	Water	NA	NA	20.1	12.01	01/20/2020	17:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY L1210

WorkList ID: 135197

Department: Wet-Chemistry

Date: 01-20-2020 15:02:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/20/2020	Solid	L1004-15	Corrosivity	Cool 4 deg C	PSEG05	I51	BU-03-012020	01/20/2020	9045D
01/20/2020	Solid	L1210-07	Corrosivity	Cool 4 deg C	ATCE02	I32	Q012-WC1	01/20/2020	9045D
01/20/2020	Solid	L1211-02	Corrosivity	Cool 4 deg C	PSEG01	I32	B1-0-15	01/20/2020	9045D
01/20/2020	Solid	L1211-04	Corrosivity	Cool 4 deg C	PSEG01	I32	B1-15-30	01/20/2020	9045D

Date/Time 01-20-2020 16:00
 Received by: Jc
 Relinquished by: Jc

Date/Time 01-20-2020 17:35
 Received by: Jc
 Relinquished by: Jc