



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107275

Slope : 98.0

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/21/2020	10:25
2	CAL2	1	Water	NA	NA	20.1	7.01	01/21/2020	10:26
3	CAL3	1	Water	NA	NA	20.1	10.04	01/21/2020	10:27
4	ICV	1	Water	NA	NA	20.2	6.99	01/21/2020	10:30
5	CCV1	1	Water	NA	NA	20.2	2.01	01/21/2020	10:31
6	L1212-01	1	Water	NA	NA	11.7	4.43	01/21/2020	10:33
7	L1212-02	1	Water	NA	NA	13.7	4.54	01/21/2020	10:35
8	L1212-03	1	Water	NA	NA	15.2	4.58	01/21/2020	10:36
9	L1212-03DUP	1	Water	NA	NA	15.1	4.59	01/21/2020	10:37
10	CCV2	1	Water	NA	NA	20.2	12.02	01/21/2020	10:40

107275

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1212 WorkList ID : 135225 Department : Wet-Chemistry Date : 01-21-2020 10:16:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/20/2020	Water	L1212-01 -1	pH	Cool 4 deg C	STER09	I51	CELL-A-POOL-WATER-(DE-I)	01/20/2020	SM 4500-H B
01/20/2020	Water	L1212-02 -1	pH	Cool 4 deg C	STER09	I51	CELL-B-C-POOL-WATER-(DE-I)	01/20/2020	SM 4500-H B
01/20/2020	Water	L1212-03 -1	pH	Cool 4 deg C	STER09	I51	CELL-D-POOL-WATER-(DE-I)	01/20/2020	SM 4500-H B

Date/Time 01-21-2020 10-20
 Received by: JP
 Relinquished by: AV

Date/Time 01-21-2020 11-10
 Received by: AV
 Relinquished by: AV