



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107276

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP81377
Turbidity Calibration std, 0NTU	WP81375
Turbidity Calibration std, 20NTU	WP81373
Turbidity Calibration std, 80NTU	WP81379
Turbidity Calibration std, 40NTU	WP81372
Turbidity Calibration std, 1NTU	WP81378
Turbidity Calibration std, 10NTU	WP81374
Turbidity Calibration std, 5NTU	WP81376
10 NTU Standard 500 ml	W2489

Intercept: -0.4058

Slope: 0.958

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.162	01/21/2020	11:17
2	CAL2	1	1	0.996	01/21/2020	11:09
3	CAL3	5	1	4.370	01/21/2020	11:22
4	CAL4	10	1	9.380	01/21/2020	11:25
5	CAL5	20	1	17.400	01/21/2020	11:28
6	CAL6	40	1	37.600	01/21/2020	11:31
7	CAL7	80	1	76.700	01/21/2020	11:34



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB107276

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	10.027	01/21/2020	11:37
2	ICB		1	0.155	0.585	01/21/2020	11:40
3	CCV1	10	1	9.280	10.110	01/21/2020	11:43
4	CCB1		1	0.158	0.589	01/21/2020	11:46
5	LB107276BL		1	0.149	0.579	01/21/2020	11:49
6	L1173-04		1	50.500	53.138	01/21/2020	11:53
7	L1173-04DUP		1	53.200	55.956	01/21/2020	11:56
8	CCV2	10	1	9.270	10.100	01/21/2020	11:59
9	CCB2		1	0.156	0.586	01/21/2020	12:03

WORKLIST(Hardcopy Internal Chain)

C13107276

WorkList Name : TUR01212020 WorkList ID : 135226 Department : Wet-Chemistry Date : 01-21-2020 10:30:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/17/2020	Water	L1173-04	Turbidity	Cool 4 deg C	AQUA04	I31	0051-DECANT-H2O-COMP-3	01/15/2020	SM2130 B

Date/Time 01-21-2020 10:45
 Received by: AK
 Relinquished by: SP

Date/Time 01-21-2020 12:25
 Received by: SR
 Relinquished by: AK