


LB 107282

ISM02.4

Instrument ID = CV1

μ Abs.:

40714



Linear

A= 0.0000e+000

B= 2.4702e-004

Slope

C= -2.5702e-002

Intercept

Rho= 0.9999603

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% ID
0.00	0.000	0.026	0.026	210	0.000	210					
0.05	0.050										
0.20	0.200	0.219	0.019	990	0.0 %	990					100
2.50	2.500	2.482	-0.018	10153	0.0 %	10153					-1
5.00	5.000	4.936	-0.064	20086	0.0 %	20086					-1
7.50	7.500	7.506	0.006	30489	0.0 %	30489					0
10.0	10.000	10.031	0.031	40714	0.0 %	40714					0

LB107282

INSTRUMENT ID :CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	210	-		0 ISM02.4	PPB	1/21/2020 15:39	Std
0.2 S0.2	990	-		0.2 ISM02.4	PPB	1/21/2020 15:42	Std
2.5 S2.5	10153	-		2.5 ISM02.4	PPB	1/21/2020 15:44	Std
5 S5.0	20086	-		5 ISM02.4	PPB	1/21/2020 15:46	Std
7.5 S7.5	30489	-		7.5 ISM02.4	PPB	1/21/2020 15:48	Std
10 S10.0	40714	-		10 ISM02.4	PPB	1/21/2020 15:51	Std
ICV	ICV	16282	3.9962 -	ISM02.4	PPB	1/21/2020 15:54	SMPL
ICB	ICB	166	0.0153 -	ISM02.4	PPB	1/21/2020 15:56	SMPL
CCV085	CCV085	22283	5.4786 -	ISM02.4	PPB	1/21/2020 15:58	SMPL
CCB085	CCB085	157	0.0131 -	ISM02.4	PPB	1/21/2020 16:00	SMPL
PB126249E	PBS249	227	0.0304 -	ISM02.4	PPB	1/21/2020 16:02	SMPL
L1137-01	MF3E04	526	0.1042 -	ISM02.4	PPB	1/21/2020 16:05	SMPL
L1137-02	MF3E04D	508	0.0998 -	ISM02.4	PPB	1/21/2020 16:07	SMPL
L1137-04	MF3E63	124871	30.8194 -	ISM02.4	PPB	1/21/2020 16:14	SMPL
L1137-05	MF3E64	16824	4.1301 -	ISM02.4	PPB	1/21/2020 16:16	SMPL
L1137-06	MF3E67	3061	0.7304 -	ISM02.4	PPB	1/21/2020 16:18	SMPL
L1137-07	MF3E70	2511	0.5946 -	ISM02.4	PPB	1/21/2020 16:21	SMPL
L1137-08	MF3E71	26785	6.5906 -	ISM02.4	PPB	1/21/2020 16:23	SMPL
L1137-09	MF3E72	9193	2.2451 -	ISM02.4	PPB	1/21/2020 16:25	SMPL
L1137-10	MF3E73	596	0.1215 -	ISM02.4	PPB	1/21/2020 16:27	SMPL
L1137-11	MF3E60	1445	0.3312 -	ISM02.4	PPB	1/21/2020 16:30	SMPL
L1137-12	MF3E62	952	0.2095 -	ISM02.4	PPB	1/21/2020 16:32	SMPL
L1137-13	MF3E65	1013	0.2245 -	ISM02.4	PPB	1/21/2020 16:34	SMPL
L1137-14	MF3E66	1508	0.3468 -	ISM02.4	PPB	1/21/2020 16:36	SMPL
L1137-15	MF3E68	4391	1.0589 -	ISM02.4	PPB	1/21/2020 16:39	SMPL
L1137-16	MF3E69	4469	1.0782 -	ISM02.4	PPB	1/21/2020 16:41	SMPL
L1137-17	MF3E78	1923	0.4493 -	ISM02.4	PPB	1/21/2020 16:43	SMPL
L1137-18	MF3E83	9503	2.3217 -	ISM02.4	PPB	1/21/2020 16:45	SMPL
L1137-19	MF3E84	2415	0.5708 -	ISM02.4	PPB	1/21/2020 16:48	SMPL
CCV086	CCV086	21198	5.2105 -	ISM02.4	PPB	1/21/2020 16:50	SMPL
CCB086	CCB086	170	0.0163 -	ISM02.4	PPB	1/21/2020 16:52	SMPL
L1137-20	MF3E85	0	-0.0257 -	ISM02.4	PPB	1/21/2020 16:57	SMPL
L1137-03	MF3E04S	10598	2.5922 -	ISM02.4	PPB	1/21/2020 16:59	SMPL
L1137-20	MF3E85	11161	2.7312 -	ISM02.4	PPB	1/21/2020 17:01	SMPL
L1137-21	MF3E86	1766	0.4105 -	ISM02.4	PPB	1/21/2020 17:03	SMPL
PB126250E	PBS250	201	0.0239 -	ISM02.4	PPB	1/21/2020 17:05	SMPL
L1158-01	MF3E46	3487	0.8356 -	ISM02.4	PPB	1/21/2020 17:08	SMPL
L1158-02	MF3E46D	3463	0.8297 -	ISM02.4	PPB	1/21/2020 17:10	SMPL
L1158-03	MF3E46S	13669	3.3508 -	ISM02.4	PPB	1/21/2020 17:12	SMPL
L1158-04	MF3E54	2074	0.4866 -	ISM02.4	PPB	1/21/2020 17:14	SMPL
L1158-05	MF3E56	7464	1.818 -	ISM02.4	PPB	1/21/2020 17:17	SMPL
L1158-06	MF3E57	2923	0.6963 -	ISM02.4	PPB	1/21/2020 17:19	SMPL
L1158-07	MF3E58	866	0.1882 -	ISM02.4	PPB	1/21/2020 17:21	SMPL
L1158-08	MF3E59	502	0.0983 -	ISM02.4	PPB	1/21/2020 17:24	SMPL

L1158-09	MF3E61	595	0.1213 -	ISM02.4	PPB	1/21/2020 17:26 SMPL
PB126251E	PBS251	207	0.0254 -	ISM02.4	PPB	1/21/2020 17:28 SMPL
L1157-01	MF3E05	5140	1.244 -	ISM02.4	PPB	1/21/2020 17:30 SMPL
L1157-02	MF3E06	1145	0.2571 -	ISM02.4	PPB	1/21/2020 17:33 SMPL
L1157-03	MF3E07	537	0.1069 -	ISM02.4	PPB	1/21/2020 17:35 SMPL
L1157-04	MF3E08	826	0.1783 -	ISM02.4	PPB	1/21/2020 17:37 SMPL
L1157-05	MF3E09	3568	0.8557 -	ISM02.4	PPB	1/21/2020 17:39 SMPL
CCV087	CCV087	21444	5.2713 -	ISM02.4	PPB	1/21/2020 17:42 SMPL
CCB087	CCB087	166	0.0153 -	ISM02.4	PPB	1/21/2020 17:44 SMPL
L1157-06	MF3E40	426293	HIGH -	ISM02.4	PPB	1/21/2020 17:46
L1157-07	MF3E41	10165	2.4852 -	ISM02.4	PPB	1/21/2020 17:48 SMPL
L1157-08	MF3E42	1608	0.3715 -	ISM02.4	PPB	1/21/2020 17:50 SMPL
L1157-09	MF3E43	749	0.1593 -	ISM02.4	PPB	1/21/2020 17:53 SMPL
L1157-10	MF3E44	4297	1.0357 -	ISM02.4	PPB	1/21/2020 17:55 SMPL
L1157-11	MF3E50	356561	HIGH -	ISM02.4	PPB	1/21/2020 17:57
L1157-13	MF3E51D	16094	3.9498 -	ISM02.4	PPB	1/21/2020 18:01 SMPL
L1157-14	MF3E51S	25748	6.3345 -	ISM02.4	PPB	1/21/2020 18:04 SMPL
L1157-15	MF3E52	8284	2.0206 -	ISM02.4	PPB	1/21/2020 18:06 SMPL
L1157-16	MF3E53	3072	0.7331 -	ISM02.4	PPB	1/21/2020 18:08 SMPL
L1157-17	MF3E55	1033	0.2295 -	ISM02.4	PPB	1/21/2020 18:10 SMPL
L1157-18	MF3E77	633	0.1307 -	ISM02.4	PPB	1/21/2020 18:13 SMPL
L1157-19	MF3E45	51900	12.7944 -	ISM02.4	PPB	1/21/2020 18:15 SMPL
L1157-20	MF3E47	565	0.1139 -	ISM02.4	PPB	1/21/2020 18:17 SMPL
L1157-21	MF3E48	721	0.1524 -	ISM02.4	PPB	1/21/2020 18:19 SMPL
L1157-22	MF3E49	943	0.2072 -	ISM02.4	PPB	1/21/2020 18:22 SMPL
L1157-06D	MF3E40	31171	7.674 -	ISM02.4	PPB	1/21/2020 18:24 SMPL
L1157-11D	MF3E50	47907	11.8081 -	ISM02.4	PPB	1/21/2020 18:26 SMPL
CCV088	CCV088	21208	5.213 -	ISM02.4	PPB	1/21/2020 18:31 SMPL
CCB088	CCB088	38	-0.0163 -	ISM02.4	PPB	1/21/2020 18:33 SMPL
L1157-19D	MF3E45	25311	6.2265 -	ISM02.4	PPB	1/21/2020 18:35 SMPL
L1157-11D	MF3E50	24767	6.0921 -	ISM02.4	PPB	1/21/2020 18:38 SMPL
L1157-12	MF3E51	15166	3.7205 -	ISM02.4	PPB	1/21/2020 18:40 SMPL
CCV089	CCV089	21546	5.2965 -	ISM02.4	PPB	1/21/2020 18:42 SMPL
CCB089	CCB089	39	-0.0161 -	ISM02.4	PPB	1/21/2020 18:44 SMPL