



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107283

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/21/2020	15:40
2	CAL2	1	Water	NA	NA	20.1	7.01	01/21/2020	15:42
3	CAL3	1	Water	NA	NA	20.1	10.03	01/21/2020	15:45
4	ICV	1	Water	NA	NA	20.2	7.00	01/21/2020	15:47
5	CCV1	1	Water	NA	NA	20.1	2.01	01/21/2020	15:48
6	L1216-09	1	Solid	20.02	20	20.0	6.30	01/21/2020	15:50
7	L1216-09DUP	1	Solid	20.03	20	21.0	6.31	01/21/2020	15:52
8	L1216-10	1	Solid	20.05	20	20.6	6.86	01/21/2020	15:53
9	L1216-11	1	Solid	20.04	20	20.2	7.02	01/21/2020	15:55
10	L1216-12	1	Solid	20.03	20	19.9	6.48	01/21/2020	15:56
11	CCV2	1	Water	NA	NA	20.1	12.01	01/21/2020	15:57