



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107295

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/22/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/22/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/22/2020	16:02
4	ICV	1	Water	NA	NA	20.1	7.00	01/22/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/22/2020	16:06
6	L1222-03	1	Water	NA	NA	14.5	7.14	01/22/2020	16:07
7	L1222-03DUP	1	Water	NA	NA	14.6	7.15	01/22/2020	16:08
8	CCV2	1	Water	NA	NA	20.1	12.02	01/22/2020	16:10

283167295

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1230 WorkList ID : 135267 Department : Wet-Chemistry Date : 01-22-2020 15:47:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2020	Water	L1222-03 - 13	Corrosivity	Cool 4 deg C	DVIR01	I53	SH-BC	01/21/2020	9040C

Date/Time 01-22-2020 15:50
Received by: JD
Relinquished by: JD

Date/Time 01-22-2020 17:10
Received by: JD
Relinquished by: JD