



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/23/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 01/22/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/23/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107297

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1219-01	ST-CONCRETE	1	1.00	2.00	2.00	100.0	Concrete Sample ,100% solids
L1225-01	PF-5	2	1.14	9.76	8.1	80.7	
L1226-01	B-1	3	1.15	9.98	9.81	98.1	
L1226-02	B-2	4	1.15	9.63	8.75	89.6	
L1226-03	B-3	5	1.16	9.96	9.04	89.5	
L1226-04	B-4	6	1.13	9.62	8.74	89.6	
L1226-05	B-5	7	1.17	9.77	8.71	87.7	
L1226-06	B-6	8	1.18	9.89	8.86	88.2	
L1230-01	STOCK-PILE-A	9	1.13	9.9	7.82	76.3	
L1230-03	STOCK-PILE-B	10	1.16	9.97	8.58	84.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$