


U3107309
INSTRUMENT ID : CV1

ISM02.4



Linear

A= 0.0000e+000

B= 2.6077e-004

Slope

C= -9.2933e-002

Intercept

Rho= 0.9999393

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.024	-0.024	266	0.000	266					
0.05	0.050										
0.20	0.200	0.183	-0.017	1059	0.0 %	1059					-9
2.50	2.500	2.508	0.008	9973	0.0 %	9973					0
5.00	5.000	5.043	0.043	19697	0.0 %	19697					1
7.50	7.500	7.554	0.054	29325	0.0 %	29325					1
10.0	10.000	9.935	-0.065	38456	0.0 %	38456					-1

LB107309

INSTRUMENT ID :CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	266	-	0	ISM02.4	PPB	1/23/2020 11:53	Std
0.2 S0.2	1059	-	0.2	ISM02.4	PPB	1/23/2020 11:56	Std
2.5 S2.5	9973	-	2.5	ISM02.4	PPB	1/23/2020 11:58	Std
5 S5.0	19697	-	5	ISM02.4	PPB	1/23/2020 12:00	Std
7.5 S7.5	29325	-	7.5	ISM02.4	PPB	1/23/2020 12:02	Std
10 S10.0	38456	-	10	ISM02.4	PPB	1/23/2020 12:05	Std
ICV	ICV	16071	4.0979 -	ISM02.4	PPB	1/23/2020 12:14	SMPL
ICB	ICB	212	-0.0376 -	ISM02.4	PPB	1/23/2020 12:16	SMPL
CCV096	CCV096	20389	5.2239 -	ISM02.4	PPB	1/23/2020 12:18	SMPL
CCB096	CCB096	209	-0.0384 -	ISM02.4	PPB	1/23/2020 12:20	SMPL
PB126258E	PBS258	33	-0.0843 -	ISM02.4	PPB	1/23/2020 12:23	SMPL
L1119-01	MF3E20	4297	1.0276 -	ISM02.4	PPB	1/23/2020 12:25	SMPL
L1119-02	MF3E20D	4332	1.0367 -	ISM02.4	PPB	1/23/2020 12:27	SMPL
L1119-03	MF3E20S	14035	3.5669 -	ISM02.4	PPB	1/23/2020 12:29	SMPL
L1119-04	MF3E21	4643	1.1178 -	ISM02.4	PPB	1/23/2020 12:32	SMPL
L1119-05	MF3E22	803	0.1165 -	ISM02.4	PPB	1/23/2020 12:34	SMPL
L1119-06	MF3E23	292	-0.0168 -	ISM02.4	PPB	1/23/2020 12:36	SMPL
L1119-07	MF3E24	280	-0.0199 -	ISM02.4	PPB	1/23/2020 12:38	SMPL
L1119-08	MF3E25	12808	3.247 -	ISM02.4	PPB	1/23/2020 12:41	SMPL
L1119-09	MF3E26	1193	0.2182 -	ISM02.4	PPB	1/23/2020 12:43	SMPL
L1119-10	MF3E27	1005	0.1691 -	ISM02.4	PPB	1/23/2020 12:45	SMPL
L1119-11	MF3E28	619	0.0685 -	ISM02.4	PPB	1/23/2020 12:47	SMPL
L1119-12	MF3E29	264	-0.0241 -	ISM02.4	PPB	1/23/2020 12:50	SMPL
L1119-14	MF3E81	296	-0.0157 -	ISM02.4	PPB	1/23/2020 12:52	SMPL
L1119-15	MF3E30	944	0.1532 -	ISM02.4	PPB	1/23/2020 12:54	SMPL
L1119-16	MF3E31	959	0.1571 -	ISM02.4	PPB	1/23/2020 12:56	SMPL
L1119-17	MF3E32	18215	4.657 -	ISM02.4	PPB	1/23/2020 12:59	SMPL
L1119-18	MF3E33	613	0.0669 -	ISM02.4	PPB	1/23/2020 13:01	SMPL
L1119-19	MF3E34	489	0.0346 -	ISM02.4	PPB	1/23/2020 13:03	SMPL
L1119-20	MF3E36	873	0.1347 -	ISM02.4	PPB	1/23/2020 13:06	SMPL
CCV097	CCV097	19675	5.0377 -	ISM02.4	PPB	1/23/2020 13:16	SMPL
CCB097	CCB097	121	-0.0614 -	ISM02.4	PPB	1/23/2020 13:18	SMPL
L1119-21	MF3E37	846	0.1277 -	ISM02.4	PPB	1/23/2020 13:20	SMPL
L1119-22	MF3E38	571	0.056 -	ISM02.4	PPB	1/23/2020 13:23	SMPL
PB126259E	PBS259	-6	-0.0945 -	ISM02.4	PPB	1/23/2020 13:25	SMPL
L1120-01	MF3E35	916	0.1459 -	ISM02.4	PPB	1/23/2020 13:27	SMPL
L1120-02	MF3E35D	913	0.1451 -	ISM02.4	PPB	1/23/2020 13:29	SMPL
L1120-03	MF3E35S	10466	2.6363 -	ISM02.4	PPB	1/23/2020 13:32	SMPL
L1120-04	MF3E79	5146	1.249 -	ISM02.4	PPB	1/23/2020 13:34	SMPL
L1120-05	MF3E80	4097	0.9754 -	ISM02.4	PPB	1/23/2020 13:36	SMPL
L1120-06	MF3E82	1381	0.2672 -	ISM02.4	PPB	1/23/2020 13:38	SMPL
L1120-07	MF3E10	863	0.1321 -	ISM02.4	PPB	1/23/2020 13:41	SMPL
L1120-08	MF3E11	1113	0.1973 -	ISM02.4	PPB	1/23/2020 13:43	SMPL
L1120-09	MF3E12	898	0.1412 -	ISM02.4	PPB	1/23/2020 13:45	SMPL

L1120-10	MF3E13	343	-0.0035 -	ISM02.4	PPB	1/23/2020 13:47 SMPL
L1120-11	MF3E14	570	0.0557 -	ISM02.4	PPB	1/23/2020 13:50 SMPL
L1120-12	MF3E15	1207	0.2218 -	ISM02.4	PPB	1/23/2020 13:52 SMPL
L1120-13	MF3E16	566	0.0547 -	ISM02.4	PPB	1/23/2020 13:54 SMPL
L1120-14	MF3E17	711	0.0925 -	ISM02.4	PPB	1/23/2020 13:56 SMPL
L1120-15	MF3E18	366	0.0025 -	ISM02.4	PPB	1/23/2020 13:59 SMPL
L1120-16	MF3E19	551	0.0508 -	ISM02.4	PPB	1/23/2020 14:01 SMPL
L1120-17	MF3E39	701	0.0899 -	ISM02.4	PPB	1/23/2020 14:03 SMPL
CCV098	CCV098	19880	5.0911 -	ISM02.4	PPB	1/23/2020 14:05 SMPL
CCB098	CCB098	43	-0.0817 -	ISM02.4	PPB	1/23/2020 14:08 SMPL
L1120-18	MF3E00	2785	0.6333 -	ISM02.4	PPB	1/23/2020 14:10 SMPL
L1120-19	MF3E01	526	0.0442 -	ISM02.4	PPB	1/23/2020 14:12 SMPL
L1120-20	MF3E02	424	0.0176 -	ISM02.4	PPB	1/23/2020 14:15 SMPL
L1120-21	MF3E03	290	-0.0173 -	ISM02.4	PPB	1/23/2020 14:17 SMPL
L1137-04D	MF3E63	24578	6.3162 -	ISM02.4	PPB	1/23/2020 14:19 SMPL
CCV099	CCV099	19658	5.0332 -	ISM02.4	PPB	1/23/2020 14:21 SMPL
CCB099	CCB099	13	-0.0895 -	ISM02.4	PPB	1/23/2020 14:24 SMPL