



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107312

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/23/2020	12:45
2	CAL2	1	Water	NA	NA	20.1	7.01	01/23/2020	12:46
3	CAL3	1	Water	NA	NA	20.1	10.03	01/23/2020	12:47
4	ICV	1	Water	NA	NA	20.1	7.00	01/23/2020	12:48
5	CCV1	1	Water	NA	NA	20.2	2.01	01/23/2020	13:10
6	L1236-03	1	Water	NA	NA	12.4	6.85	01/23/2020	13:11
7	L1236-03DUP	1	Water	NA	NA	12.5	6.86	01/23/2020	13:12
8	L1236-04	1	Water	NA	NA	12.1	5.53	01/23/2020	13:15
9	L1236-05	1	Water	NA	NA	12.4	7.01	01/23/2020	13:16
10	L1236-06	1	Water	NA	NA	13.1	5.82	01/23/2020	13:20
11	L1236-09	1	Water	NA	NA	13.2	7.07	01/23/2020	13:21
12	L1236-10	1	Water	NA	NA	13.4	6.97	01/23/2020	13:22
13	L1236-11	1	Water	NA	NA	14.0	6.87	01/23/2020	13:25
14	L1236-12	1	Water	NA	NA	14.6	7.01	01/23/2020	13:26
15	L1236-13	1	Water	NA	NA	14.2	6.78	01/23/2020	13:30
16	CCV2	1	Water	NA	NA	20.1	12.02	01/23/2020	13:32
17	L1236-14	1	Water	NA	NA	14.1	6.86	01/23/2020	13:35
18	L1236-15	1	Water	NA	NA	15.5	8.16	01/23/2020	13:36
19	CCV3	1	Water	NA	NA	20.1	2.01	01/23/2020	13:40

HB 101312

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W/L1236 WorkList ID : 135309 Department : Wet-Chemistry Date : 01-23-2020 12:38:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2020	Water	L1236-03 ✓	pH	Cool 4 deg C	PSEG01	I41	40653	01/23/2020	9040C
01/23/2020	Water	L1236-04 ✓	pH	Cool 4 deg C	PSEG01	I41	40655	01/23/2020	9040C
01/23/2020	Water	L1236-05 ✓	pH	Cool 4 deg C	PSEG01	I41	40660	01/23/2020	9040C
01/23/2020	Water	L1236-06 ✓	pH	Cool 4 deg C	PSEG01	I41	40658	01/23/2020	9040C
01/23/2020	Water	L1236-09 ✓	pH	Cool 4 deg C	PSEG01	I41	30636	01/23/2020	9040C
01/23/2020	Water	L1236-10 ✓	pH	Cool 4 deg C	PSEG01	I41	30637	01/23/2020	9040C
01/23/2020	Water	L1236-11 ✓	pH	Cool 4 deg C	PSEG01	I41	30638	01/23/2020	9040C
01/23/2020	Water	L1236-12 ✓	pH	Cool 4 deg C	PSEG01	I41	30639	01/23/2020	9040C
01/23/2020	Water	L1236-13 ✓	pH	Cool 4 deg C	PSEG01	I41	40507	01/23/2020	9040C
01/23/2020	Water	L1236-14 ✓	pH	Cool 4 deg C	PSEG01	I41	40508	01/23/2020	9040C
01/23/2020	Water	L1236-15- 1 ✓	pH	Cool 4 deg C	PSEG01	I41	40218-W	01/23/2020	9040C

Date/Time 01-23-2020

Received by: JP

Relinquished by: JP

Date/Time 01-23-2020

Received by: JP

Relinquished by: JP