


7470A

LB 107319
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38236



A= 0.0000e+000

B= 2.6058e-004

C= -8.1666e-002

Rho= 0.9997870

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.072	-0.072	37	0.000	37				
0.2	0.200	0.186	-0.014	1028	0.0 %	1028				
2.5	2.500	2.548	0.048	10091	0.0 %	10091				
5.0	5.000	5.056	0.056	19718	0.0 %	19718				
7.5	7.500	7.600	0.100	29479	0.0 %	29479				
10.0	10.000	9.882	-0.118	38236	0.0 %	38236				

LB107319 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jan 2020 16:36:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD.0	37	-	0.0000	7470A	PPB	23 Jan 2020 16:36:44	S	Std
0.2 - 1	SD.2	1028	-	0.2000	7470A	PPB	23 Jan 2020 16:38:52	S	Std
2.5 - 1	SD.5	10091	-	2.5000	7470A	PPB	23 Jan 2020 16:40:59	S	Std
5.0 - 1	SD.5	19718	-	5.0000	7470A	PPB	23 Jan 2020 16:43:06	S	Std
7.5 - 1	SD.5	29479	-	7.5000	7470A	PPB	23 Jan 2020 16:45:14	S	Std
10.0 - 1	SD.5	38236	-	10.0000	7470A	PPB	23 Jan 2020 16:47:20	S	Std
ICV84 - 1	ICV84	15682	4.0047	-	7470A	PPB	23 Jan 2020 16:51:22	U	SMPL
ICB84 - 1	ICB84	1	-0.0814	-	7470A	PPB	23 Jan 2020 16:53:27	U	SMPL
CCV04 - 1	CCV04	19801	5.0780	-	7470A	PPB	23 Jan 2020 16:55:32	U	SMPL
CCB04 - 1	CCB04	26	-0.0749	-	7470A	PPB	23 Jan 2020 16:57:38	U	SMPL
CRA - 1	CRA	1140	0.2154	-	7470A	PPB	23 Jan 2020 16:59:44	U	SMPL
HighStd - 1	HighStd	38021	9.8257	-	7470A	PPB	23 Jan 2020 17:01:49	U	SMPL
ChkStd - 1	ChkStd	27236	7.0154	-	7470A	PPB	23 Jan 2020 17:03:56	U	SMPL
PB126272BL - 1	PBW	-21	-0.0871	-	7470A	PPB	23 Jan 2020 17:06:02	U	SMPL
PB126272BS - 1	LCSW	15404	3.9323	-	7470A	PPB	23 Jan 2020 17:08:09	U	SMPL
L1004-15 - 1	BU-03-012020	172	-0.0368	-	7470A	PPB	23 Jan 2020 17:10:16	U	SMPL
L1004-15DUP - 1	BU-03-012020DUP	184	-0.0337	-	7470A	PPB	23 Jan 2020 17:12:23	U	SMPL
L1004-15MS - 1	BU-03-012020MS	16712	4.2731	-	7470A	PPB	23 Jan 2020 17:14:31	U	SMPL
L1004-15MSD - 1	BU-03-012020MSD	17262	4.4164	-	7470A	PPB	23 Jan 2020 17:16:39	U	SMPL
L1154-02 - 1	B4-0-15	170	-0.0374	-	7470A	PPB	23 Jan 2020 17:18:45	U	SMPL
CCV05 - 1	CCV05	20406	5.2357	-	7470A	PPB	23 Jan 2020 17:20:51	U	SMPL
CCB05 - 1	CCB05	0	-0.0817	-	7470A	PPB	23 Jan 2020 17:22:57	U	SMPL
L1154-04 - 1	B4-15-20	169	-0.0376	-	7470A	PPB	23 Jan 2020 17:25:02	U	SMPL
L1154-06 - 1	B3-0-15	171	-0.0371	-	7470A	PPB	23 Jan 2020 17:27:08	U	SMPL
L1154-08 - 1	B3-15-20	171	-0.0371	-	7470A	PPB	23 Jan 2020 17:29:13	U	SMPL
L1155-03 - 1	RBR-200028	172	-0.0368	-	7470A	PPB	23 Jan 2020 17:31:19	U	SMPL
L1155-06 - 1	RBR-200052	177	-0.0355	-	7470A	PPB	23 Jan 2020 17:33:26	U	SMPL
L1210-07 - 1	Q012-WC1	177	-0.0355	-	7470A	PPB	23 Jan 2020 17:35:32	U	SMPL
L1211-02 - 1	B1-0-15	157	-0.0408	-	7470A	PPB	23 Jan 2020 17:37:39	U	SMPL
L1211-04 - 1	B1-15-30	165	-0.0387	-	7470A	PPB	23 Jan 2020 17:39:46	U	SMPL
PB126273BL - 1	PBW	8	-0.0796	-	7470A	PPB	23 Jan 2020 17:44:42	U	SMPL
PB126273BS - 1	LCSW	15751	4.0227	-	7470A	PPB	23 Jan 2020 17:46:48	U	SMPL
CCV06 - 1	CCV06	19466	4.9907	-	7470A	PPB	23 Jan 2020 17:58:09	U	SMPL
CCB06 - 1	CCB06	22	-0.0759	-	7470A	PPB	23 Jan 2020 18:00:15	U	SMPL
L1213-02 - 1	TP-1	180	-0.0348	-	7470A	PPB	23 Jan 2020 18:02:21	U	SMPL
L1213-02DUP - 1	TP-1DUP	168	-0.0379	-	7470A	PPB	23 Jan 2020 18:04:27	U	SMPL
L1213-02MS - 1	TP-1MS	17845	4.5683	-	7470A	PPB	23 Jan 2020 18:06:34	U	SMPL
L1213-02MSD - 1	TP-1MSD	17316	4.4305	-	7470A	PPB	23 Jan 2020 18:08:40	U	SMPL
L1213-04 - 1	TP-2	174	-0.0363	-	7470A	PPB	23 Jan 2020 18:10:46	U	SMPL
L1216-09 - 1	SB-01-COMP	170	-0.0374	-	7470A	PPB	23 Jan 2020 18:12:53	U	SMPL
L1216-10 - 1	SB-02-COMP	194	-0.0311	-	7470A	PPB	23 Jan 2020 18:15:00	U	SMPL
L1216-11 - 1	SB-03-COMP	172	-0.0368	-	7470A	PPB	23 Jan 2020 18:17:07	U	SMPL
L1216-12 - 1	SB-04-COMP	170	-0.0374	-	7470A	PPB	23 Jan 2020 18:19:15	U	SMPL
PB126272TB - 1	PB126226TB	11	-0.0788	-	7470A	PPB	23 Jan 2020 18:21:21	U	SMPL
CCV07 - 1	CCV07	19352	4.9610	-	7470A	PPB	23 Jan 2020 18:23:27	U	SMPL
CCB07 - 1	CCB07	7	-0.0798	-	7470A	PPB	23 Jan 2020 18:25:34	U	SMPL
PB126273TB - 1	PB126261TB	10	-0.0791	-	7470A	PPB	23 Jan 2020 18:27:40	U	SMPL
L1004-15LX5 - 1		29	-0.0741	-	7470A	PPB	23 Jan 2020 18:29:46	U	SMPL
L1213-02LX5 - 1		41	-0.0710	-	7470A	PPB	23 Jan 2020 18:31:52	U	SMPL
CCV08 - 1	CCV08	20026	5.1366	-	7470A	PPB	23 Jan 2020 18:33:58	U	SMPL
CCB08 - 1	CCB08	12	-0.0785	-	7470A	PPB	23 Jan 2020 18:36:04	U	SMPL