

UB107323


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

38845



A= 0.0000e+000

B= 2.5840e-004

C= -2.8795e-002

Rho= 0.9999689

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	116	0.000	116				
0.05	0.050	0.055	0.005	326	0.0 %	326				
0.2	0.200	0.226	0.026	986	0.0 %	986				
2.5	2.500	2.490	-0.010	9746	0.0 %	9746				
5.0	5.000	4.936	-0.064	19215	0.0 %	19215				
7.5	7.500	7.533	0.033	29262	0.0 %	29262				
10.0	10.000	10.009	0.009	38845	0.0 %	38845				

LB107323 INSTRUMENT ID:CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Jan 2020 10:47:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0.0	116	-	0.0000	245.1	PPB	24 Jan 2020 10:59:34	S	Std
0.05 - 1	S0.05	326	-	0.0500	245.1	PPB	24 Jan 2020 11:01:51	S	Std
0.2 - 1	S0.2	986	-	0.2000	245.1	PPB	24 Jan 2020 11:04:08	S	Std
2.5 - 1	S2.5	9746	-	2.5000	245.1	PPB	24 Jan 2020 11:06:25	S	Std
5.0 - 1	S5.0	19215	-	5.0000	245.1	PPB	24 Jan 2020 11:08:41	S	Std
7.5 - 1	S7.5	29262	-	7.5000	245.1	PPB	24 Jan 2020 11:10:56	S	Std
10.0 - 1	S10.0	38845	-	10.0000	245.1	PPB	24 Jan 2020 11:13:11	S	Std
ICV85 - 1	ICV85	16119	4.1364	-245.1		PPB	24 Jan 2020 11:42:12	U	SMPL
ICB85 - 1	ICB85	108	-0.0009	-245.1		PPB	24 Jan 2020 11:44:26	U	SMPL
CCV09 - 1	CCV09	20240	5.2013	-245.1		PPB	24 Jan 2020 11:46:41	U	SMPL
CCB09 - 1	CCB09	71	-0.0104	-245.1		PPB	24 Jan 2020 11:48:56	U	SMPL
CRA - 1	CRA	316	0.0529	-245.1		PPB	24 Jan 2020 11:51:11	U	SMPL
HighStd - 1	HighStd	40302	10.3853	-245.1		PPB	24 Jan 2020 11:53:25	U	SMPL
ChkStd - 1	ChkStd	27854	7.1687	-245.1		PPB	24 Jan 2020 11:55:41	U	SMPL
PB126312BL - 1	PBW	22	-0.0231	-245.1		PPB	24 Jan 2020 11:57:56	U	SMPL
PB126312BS - 1	LCSW	16697	4.2857	-245.1		PPB	24 Jan 2020 12:00:12	U	SMPL
L1142-02 - 1	WS-1	504	0.1014	-245.1		PPB	24 Jan 2020 12:02:28	U	SMPL
L1233-01 - 1	55-ST-25	90	-0.0055	-245.1		PPB	24 Jan 2020 12:04:44	U	SMPL
L1233-01DUP - 1	55-ST-25DUP	106	-0.0014	-245.1		PPB	24 Jan 2020 12:07:02	U	SMPL
L1233-01MS - 1	55-ST-25MS	15372	3.9434	-245.1		PPB	24 Jan 2020 12:14:19	U	SMPL
L1233-01MSD - 1	55-ST-25MSD	15394	3.9490	-245.1		PPB	24 Jan 2020 12:16:34	U	SMPL
CCV10 - 1	CCV10	20044	5.1506	-245.1		PPB	24 Jan 2020 12:18:50	U	SMPL
CCB10 - 1	CCB10	99	-0.0032	-245.1		PPB	24 Jan 2020 12:21:04	U	SMPL
L1233-01LX5 - 1		23	-0.0229	-245.1		PPB	24 Jan 2020 12:23:19	U	SMPL
CCV11 - 1	CCV11	20451	5.2558	-245.1		PPB	24 Jan 2020 12:25:34	U	SMPL
CCB11 - 1	CCB11	41	-0.0182	-245.1		PPB	24 Jan 2020 12:27:50	U	SMPL