



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107327

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP81488
Turbidity Calibration std, 0NTU	WP81486
Turbidity Calibration std, 20NTU	WP81472
Turbidity Calibration std, 40NTU	WP81471
Turbidity Calibration std, 1NTU	WP81475
Turbidity Calibration std, 10NTU	WP81473
Turbidity Calibration std, 5NTU	WP81474
10 NTU Standard 500 ml	W2489
TURBIDITY LOD STD, 0.5NTU	WP81489
Turbidity LOQ std, 1.0NTU	WP81491

Intercept: 0.1458

Slope: 0.8672

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.162	01/24/2020	10:06
2	CAL2	1	1	1.100	01/24/2020	10:08
3	CAL3	5	1	4.510	01/24/2020	10:11
4	CAL4	10	1	8.910	01/24/2020	10:14
5	CAL5	20	1	17.100	01/24/2020	10:17
6	CAL6	40	1	35.000	01/24/2020	10:19



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB107327

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.750	9.922	01/24/2020	10:22
2	ICB		1	0.160	0.016	01/24/2020	10:24
3	CCV1	10	1	8.900	10.095	01/24/2020	10:27
4	CCB1		1	0.165	0.022	01/24/2020	10:30
5	LB107327BL		1	0.156	0.012	01/24/2020	10:37
6	L1161-10		1	0.670	0.604	01/24/2020	10:39
7	L1161-11		1	1.110	1.112	01/24/2020	10:43
8	L1173-05		1	10.200	11.594	01/24/2020	10:45
9	L1173-06		1	1.900	2.023	01/24/2020	10:47
10	L1173-06DUP		1	1.920	2.046	01/24/2020	10:50
11	CCV2	10	1	8.920	10.118	01/24/2020	10:53
12	CCB2		1	0.159	0.015	01/24/2020	10:56

WORKLIST(Hardcopy Internal Chain)

03107327

WorkList Name : Turbiditylodloq-2 WorkList ID : 135348 Department : Wet-Chemistry Date : 01-24-2020 09:30:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2020	Water	L1161-10	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOD-MDL-WATER-04-QT1-20	01/16/2020	SM2130 B
01/18/2020	Water	L1161-11	Turbidity	Cool 4 deg C	CHEM02	QA Office	LOQ-WATER-05-QT1-2020	01/16/2020	SM2130 B
01/23/2020	Water	L1173-05	Turbidity	Cool 4 deg C	AQUA04	I31	0049-DECANT-H2O-COMP-1	01/21/2020	SM2130 B
01/23/2020	Water	L1173-06	Turbidity	Cool 4 deg C	AQUA04	I31	0048-SITE-WATER	01/21/2020	SM2130 B

Date/Time 01-24-2020 9:40

Received by: AL

Relinquished by: SP

Date/Time 01-24-2020 11:20

Received by: SP

Relinquished by: AK