

LB 103741


ISM02.4

INSTRUMENT ID = CV1

Linear

 μ Abs.:

39223



A= 0.0000e+000

B= 2.5720e-004

C= -8.7360e-002

Rho= 0.9998255

Accept=Accepted

Slope

Intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.044	-0.044	168	0.000	168					
0.05	0.050										
0.20	0.200	0.176	-0.024	1024	0.0 %	1024					-12
2.50	2.500	2.548	0.048	10248	0.0 %	10248					2
5.00	5.000	5.118	0.118	20238	0.0 %	20238					2
7.50	7.500	7.401	-0.099	29116	0.0 %	29116					-1
10.0	10.000	10.001	0.001	39223	0.0 %	39223					0

U3103741

Instrument ID: CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	168	-	0	ISM02.4	PPB	1/24/2020 16:16	Std
0.2 S0.2	1024	-	0.2	ISM02.4	PPB	1/24/2020 16:18	Std
2.5 S2.5	10248	-	2.5	ISM02.4	PPB	1/24/2020 16:21	Std
5 S5.0	20238	-	5	ISM02.4	PPB	1/24/2020 16:23	Std
7.5 S7.5	29116	-	7.5	ISM02.4	PPB	1/24/2020 16:25	Std
10 S10.0	39223	-	10	ISM02.4	PPB	1/24/2020 16:27	Std
ICV	ICV	16358	4.1199	-	ISM02.4	PPB	1/24/2020 16:30 SMPL
ICB	ICB	85	-0.0655	-	ISM02.4	PPB	1/24/2020 16:33 SMPL
CCV012	CCV012	19861	5.0208	-	ISM02.4	PPB	1/24/2020 16:35 SMPL
CCB012	CCB012	94	-0.0632	-	ISM02.4	PPB	1/24/2020 16:37 SMPL
PB126352E	PBW352	143	-0.0506	-	ISM02.4	PPB	1/24/2020 16:39 SMPL
L1119-13	MF3E74	42	-0.0766	-	ISM02.4	PPB	1/24/2020 16:42 SMPL
L1120-22	MF3E75	98	-0.0622	-	ISM02.4	PPB	1/24/2020 16:44 SMPL
L1158-10	MF3E76	148	-0.0493	-	ISM02.4	PPB	1/24/2020 16:46 SMPL
CCV013	CCV013	20447	5.1716	-	ISM02.4	PPB	1/24/2020 16:48 SMPL
CCB013	CCB013	98	-0.0622	-	ISM02.4	PPB	1/24/2020 16:51 SMPL