



### Analytical Summary Report

Analysis Method: 9045D  
Parameter: Corrosivity  
Run Number: LB107349  
BalanceID: WC SC-4

Analyst By : jignesh  
Supervisor Review By : apatel  
Slope : 99.1  
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/24/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/24/2020	17:32
3	CAL3	1	Water	NA	NA	20.1	10.03	01/24/2020	17:33
4	ICV	1	Water	NA	NA	20.1	7.00	01/24/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/24/2020	17:36
6	L1008-21	1	Solid	20.03	20	20.1	8.66	01/24/2020	17:37
7	L1008-21DUP	1	Solid	20.04	20	20.2	8.67	01/24/2020	17:38
8	L1010-12	1	Solid	20.02	20	22.1	8.40	01/24/2020	17:40
9	L1017-09	1	Solid	20.04	20	20.3	8.50	01/24/2020	17:42
10	L1262-02	1	Solid	20.01	20	22.1	8.04	01/24/2020	17:45
11	L1265-02	1	Solid	20.03	20	21.8	9.95	01/24/2020	17:46
12	L1269-06	1	Solid	20.04	20	20.8	5.56	01/24/2020	17:49
13	L1269-14	1	Solid	20.02	20	21.4	5.88	01/24/2020	17:50
14	CCV2	1	Water	NA	NA	20.2	12.02	01/24/2020	17:52

28710 7349

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1245      WorkList ID : 135375      Department : Wet-Chemistry      Date : 01-24-2020 13:37:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2020	Solid	L1008-21	Corrosivity	Cool 4 deg C	PSEG05		CL-02-012420	01/24/2020	9045D
01/24/2020	Solid	L1010-12	Corrosivity	Cool 4 deg C	PSEG05		OK-01-012420	01/24/2020	9045D
01/24/2020	Solid	L1017-09	Corrosivity	Cool 4 deg C	PSEG05		TR-04-012420	01/24/2020	9045D
01/24/2020	Solid	L1262-02	Corrosivity	Cool 4 deg C	PSEG01	I41	B10(0-15)	01/24/2020	9045D
01/24/2020	Solid	L1265-02	Corrosivity	Cool 4 deg C	PARS15	I63	CTY-WC-S-SB-0203	01/24/2020	9045D
01/24/2020	Solid	L1269-06	Corrosivity	Cool 4 deg C	PSEG01	I41	TP-22-21-COMP	01/24/2020	9045D
01/24/2020	Solid	L1269-14	Corrosivity	Cool 4 deg C	PSEG01	I41	TP-25-24-COMP	01/24/2020	9045D

Date/Time 01-24-2020 16:00  
 Received by: JD  
 Relinquished by: KS

Date/Time 01-24-2020 18:30  
 Received by: KS  
 Relinquished by: JD