



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107350

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/24/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/24/2020	17:32
3	CAL3	1	Water	NA	NA	20.1	10.03	01/24/2020	17:33
4	ICV	1	Water	NA	NA	20.1	7.00	01/24/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.01	01/24/2020	18:00
6	L1265-01	1	Water	NA	NA	12.9	8.77	01/24/2020	18:01
7	L1265-01DUP	1	Water	NA	NA	12.8	8.78	01/24/2020	18:02
8	L1266-01	1	Water	NA	NA	16.5	6.32	01/24/2020	18:05
9	L1267-01	1	Water	NA	NA	16.2	6.72	01/24/2020	18:06
10	L1267-03	1	Water	NA	NA	16.9	6.58	01/24/2020	18:08
11	L1267-05	1	Water	NA	NA	17.2	8.18	01/24/2020	18:10
12	L1267-07	1	Water	NA	NA	17.2	7.85	01/24/2020	18:11
13	CCV2	1	Water	NA	NA	20.1	12.02	01/24/2020	18:13

VB10f350

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1266 WorkList ID : 135381 Department : Wet-Chemistry Date : 01-24-2020 16:44:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/23/2020	Water	L1265-01	pH	Cool 4 deg C	PARS15	I63	CTY-WC-L-01	01/23/2020	9040C
01/24/2020	Water	L1266-01	pH	Cool 4 deg C	PSEG01	I41	40652	01/24/2020	9040C
01/24/2020	Water	L1267-01	pH	Cool 4 deg C	PSEG01	I41	40328	01/24/2020	9040C
01/24/2020	Water	L1267-03	pH	Cool 4 deg C	PSEG01	I41	40329	01/24/2020	9040C
01/24/2020	Water	L1267-05	pH	Cool 4 deg C	PSEG01	I41	40670	01/24/2020	9040C
01/24/2020	Water	L1267-07	pH	Cool 4 deg C	PSEG01	I41	40505	01/24/2020	9040C

Date/Time 01-24-2020 16:56
 Received by: KS
 Relinquished by: KS

Date/Time 01-24-2020 18:00
 Received by: KS
 Relinquished by: KS