



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107351

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/24/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/24/2020	17:32
3	CAL3	1	Water	NA	NA	20.1	10.03	01/24/2020	17:33
4	ICV	1	Water	NA	NA	20.1	7.00	01/24/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/24/2020	17:36
6	L1268-01	1	Solid	20.03	20	21.5	6.39	01/24/2020	18:20
7	L1268-01DUP	1	Solid	20.04	20	21.6	6.40	01/24/2020	18:21
8	CCV2	1	Water	NA	NA	20.1	12.01	01/24/2020	18:25

28107351

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1268 WorkList ID : 135382 Department : Wet-Chemistry Date : 01-24-2020 16:47:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2020	Solid	L1268-01	pH	Cool 4 deg C	PSEG01	I41	30529	01/24/2020	9045D

Date/Time 01/24/2020 16:50
Received by: JS
Relinquished by: JS

Date/Time 01/24/2020 18:10
Received by: KS
Relinquished by: JS