



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107353

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/24/2020	15:20
2	L1222-03		1	8	100.00	>212.0	>212.0	01/24/2020	15:45
3	L1222-03DUP		1	8	100.00	>212.0	>212.0	01/24/2020	16:10
4	L1233-01		1	9	100.00	>212.0	>212.0	01/24/2020	16:35
5	L1236-01		1	8	100.00	>212.0	>212.0	01/24/2020	16:55
6	L1236-07		1	8	100.00	>212.0	>212.0	01/24/2020	17:15
7	L1236-15		1	9	100.00	>212.0	>212.0	01/24/2020	17:40
8	L1256-01		1	8	100.00	>212.0	>212.0	01/24/2020	17:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12420 WorkList ID : 135379 Department : Wet-Chemistry Date : 01-24-2020 14:57:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/27/2020	Water	L1222-03	Flash Point	Cool 4 deg C	DVIR01	I53	SH-BC	01/21/2020	1010A
02/04/2020	Water	L1233-01	Flash Point	Cool 4 deg C	DIST01	I51	55-ST-25	01/22/2020	1010A
01/30/2020	Water	L1236-01	Flash Point	Cool 4 deg C	PSEG01	I41	COMP-1	01/23/2020	1010A
01/30/2020	Water	L1236-07	Flash Point	Cool 4 deg C	PSEG01	I41	COMP-2	01/23/2020	1010A
01/30/2020	Water	L1236-15	Flash Point	Cool 4 deg C	PSEG01	I41	40218-W	01/23/2020	1010A
01/31/2020	Water	L1256-01	Flash Point	Cool 4 deg C	DIST01	J11	TANK-1	01/24/2020	1010A

Date/Time 01/24/20 15:00
 Received by: KS
 Relinquished by: JS

Date/Time 01/24/20 18:00
 Received by: JS
 Relinquished by: KS