



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/28/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 18:39
In Date: 01/27/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/28/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107360

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1186-08	PCAS-2A	15	1.14	9.7	7.72	76.9	
L1186-09	PCAS-2B	16	1.18	9.62	8.39	85.4	
L1186-10	PCAS-2C	17	1.14	9.98	7.31	69.8	
L1186-11	PCAS-2D	18	1.13	9.94	7.67	74.2	
L1186-12	PCAS-2E	19	1.12	9.92	8.69	86.0	
L1269-01	TP-22-21	1	1.13	5.89	5.56	93.1	
L1269-02	TP-22-1.0-1.5	2	1.13	9.55	8.76	90.6	
L1269-03	TP-22-7.0-7.5	3	1.18	9.62	8.28	84.1	
L1269-04	TP-22-11.5-12.0	4	1.17	9.94	8.56	84.3	
L1269-05	TP-22-21-COMP	5	1.14	9.78	8.72	87.7	
L1269-06	TP-22-21-COMP	6	1.14	9.78	8.72	87.7	
L1269-07	TP-22-21-COMP	7	1.14	9.78	8.72	87.7	
L1269-09	TP-25-24	8	1.15	6.02	4.97	78.4	
L1269-10	TP-25-1.0-1.5	9	1.18	9.64	7.92	79.7	
L1269-11	TP-25-7.0-7.5	10	1.18	9.9	8.58	84.9	
L1269-12	TP-25-11.5-12.0	11	1.18	9.97	7.32	69.9	
L1269-13	TP-25-24-COMP	12	1.18	9.97	8.6	84.4	
L1269-14	TP-25-24-COMP	13	1.18	9.97	8.6	84.4	
L1269-15	TP-25-24-COMP	14	1.18	9.97	8.6	84.4	
L1279-01	400123076.0	20	1.14	9.74	2.95	21.0	sludge sample
L1280-01	400122175.0	21	1.16	9.94	1.24	0.9	sludge sample
L1281-01	400122174.0	22	1.16	9.52	1.51	4.2	sludge sample
L1282-01	400122817.1	23	1.16	9.69	1.32	1.9	sludge sample

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$