

CB10 7378

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

1/28/2020 9:40

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.040	0.0	0.092	
ICB1	-0.140	0.0	0.002	
CCV1	245.881	0.0	0.230	
CCB1	-0.345	0.0	0.002	
PB126384BL	0.010	0.0	0.002	
PB126483BS	96.906	0.0	0.092	
L1245-01	0.052	0.0	0.003	
L1245-01DUP	-0.004	0.0	0.002	
L1245-01MS	2.512	0.0	0.005	
L1245-01MSD	2.120	0.0	0.004	
L1245-02	-0.026	0.0	0.002	
L1245-03	-0.321	0.0	0.002	
L1245-04	-0.422	0.0	0.002	
L1245-05	-0.073	0.0	0.002	
CCV2	256.169	0.0	0.240	
CCB2	-0.093	0.0	0.002	

N	16
Mean	43.642
SD	87.2669
CV%	199.96

AV

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 28 08:51:08 2020

Tue Jan 28 09:31:56 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.5175	µg/l	1/28/2020 8:51:08	
5.OPPBCN	A	Total CN	P	4.4999	µg/l	1/28/2020 8:51:09	
10PPBCN	A	Total CN	P	9.6385	µg/l	1/28/2020 8:51:10	
50PPBCN	A	Total CN	P	49.4499	µg/l	1/28/2020 8:51:11	
100PPBCN	A	Total CN	P	99.7258	µg/l	1/28/2020 8:51:12	
250PPBCN	A	Total CN	P	254.1627	µg/l	1/28/2020 8:51:13	
500PPBCN	A	Total CN	P	498.0407	µg/l	1/28/2020 8:51:14	
ICV1	S	Total CN	P	96.0399	µg/l	1/28/2020 9:27:05	
ICB1	S	Total CN	P	-0.1401	µg/l	1/28/2020 9:27:06	
CCV1	S	Total CN	P	245.8813	µg/l	1/28/2020 9:27:07	
CCB1	S	Total CN	P	-0.3448	µg/l	1/28/2020 9:27:08	
PB126384BL	S	Total CN	P	0.0096	µg/l	1/28/2020 9:27:09	
PB126384BS	S	Total CN	P	96.9061	µg/l	1/28/2020 9:27:10	
L1245-01	S	Total CN	P	0.0524	µg/l	1/28/2020 9:27:11	
L1245-01DUP	S	Total CN	P	-0.0041	µg/l	1/28/2020 9:27:12	
L1245-01MS	S	Total CN	P	2.512	µg/l	1/28/2020 9:27:13	
L1245-01MSD	S	Total CN	P	2.1202	µg/l	1/28/2020 9:27:14	
L1245-02	S	Total CN	P	-0.026	µg/l	1/28/2020 9:27:15	
L1245-03	S	Total CN	P	-0.3211	µg/l	1/28/2020 9:31:52	
L1245-04	S	Total CN	P	-0.4224	µg/l	1/28/2020 9:31:53	
L1245-05	S	Total CN	P	-0.0733	µg/l	1/28/2020 9:31:54	
CCV2	S	Total CN	P	256.1693	µg/l	1/28/2020 9:31:55	
CCB2	S	Total CN	P	-0.0929	µg/l	1/28/2020 9:31:56	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

1/28/2020 8:54


Test Total CN

Accepted 1/28/2020 8:54

Factor 1079
Bias 0.002

Coeff. of det. 0.999892

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.5175	0.0000	
2	5.0PPBCN	0.007	4.4999	5.0000	
3	10PPBCN	0.011	9.6385	10.0000	
4	50PPBCN	0.048	49.4499	50.0000	
5	100PPBCN	0.095	99.7258	100.0000	
6	250PPBCN	0.238	254.1627	250.0000	
7	500PPBCN	0.464	498.0407	500.0000	