



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107381

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/28/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.00	01/28/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.04	01/28/2020	10:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/28/2020	10:20
5	CCV1	1	Water	NA	NA	20.1	2.01	01/28/2020	10:22
6	L1278-01	1	Solid	20.03	20	19.9	7.92	01/28/2020	10:24
7	L1278-01DUP	1	Solid	20.04	20	19.8	7.93	01/28/2020	10:25
8	CCV2	1	Water	NA	NA	20.2	12.02	01/28/2020	10:26

VB 107381

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1278 WorkList ID : 135445 Department : Wet-Chemistry Date : 01-28-2020 10:07:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/24/2020	Solid	L1278-01-18	pH	Cool 4 deg C	BRO004	J12	TONNELE-AVE-01	01/24/2020	9045D

Date/Time 01-28-2020 10:10
Received by: Jg
Relinquished by: Jg

Date/Time 01-28-2020 11:30
Received by: Jg
Relinquished by: Jg