



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107382

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/28/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	7.00	01/28/2020	10:16
3	CAL3	1	Water	NA	NA	20.1	10.04	01/28/2020	10:17
4	ICV	1	Water	NA	NA	20.1	7.00	01/28/2020	10:20
5	CCV1	1	Water	NA	NA	20.1	2.01	01/28/2020	11:02
6	L1283-01	1	Water	NA	NA	11.3	6.82	01/28/2020	11:05
7	L1283-01DUP	1	Water	NA	NA	11.4	6.83	01/28/2020	11:06
8	CCV2	1	Water	NA	NA	20.2	12.02	01/28/2020	11:07

WORKLIST(Hardcopy Internal Chain)

VB 107382

WorkList Name : PH L1283 W WorkList ID : 135448 Department : Wet-Chemistry Date : 01-28-2020 10:40:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2020	Water	L1283-01	pH	Cool 4 deg C	PSEG01	I42	VB16209	01/28/2020	9040C

Date/Time 01-28-2020 10:50
 Received by: Ja
 Relinquished by: Ja

Date/Time 01-28-2020 11:35
 Received by: Dk
 Relinquished by: Ja