


7471B

LB67383  
INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

6009E



A= 0.0000e+000

B= 1.6606e-004

C= -7.0605e-002

Rho= 0.9998626

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | -0.018 | -0.018 | 314   | 0.000      | 314   |       |       |       |       |
| 0.2    | 0.200  | 0.155  | -0.045 | 1360  | 0.0 %      | 1360  |       |       |       |       |
| 2.5    | 2.500  | 2.548  | 0.048  | 15766 | 0.0 %      | 15766 |       |       |       |       |
| 5.0    | 5.000  | 5.009  | 0.009  | 30589 | 0.0 %      | 30589 |       |       |       |       |
| 7.5    | 7.500  | 7.598  | 0.098  | 46176 | 0.0 %      | 46176 |       |       |       |       |
| 10.0   | 10.000 | 9.909  | -0.091 | 60096 | 0.0 %      | 60096 |       |       |       |       |

## LB107383 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Jan 2020 11:55:37

| Sample ID       | Extended ID         | $\mu$ Abs. | Conc.   | Stnd Conc | Method | Units | Date                 | Type | Type |
|-----------------|---------------------|------------|---------|-----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | SD.0                | 314        | -       | 0.0000    | 7471B  | PPB   | 28 Jan 2020 12:30:57 | S    | Std  |
| 0.2 - 1         | SD.2                | 1360       | -       | 0.2000    | 7471B  | PPB   | 28 Jan 2020 12:33:14 | S    | Std  |
| 2.5 - 1         | SD.5                | 15766      | -       | 2.5000    | 7471B  | PPB   | 28 Jan 2020 12:35:31 | S    | Std  |
| 5.0 - 1         | SD.0                | 30589      | -       | 5.0000    | 7471B  | PPB   | 28 Jan 2020 12:37:47 | S    | Std  |
| 7.5 - 1         | SD.5                | 46176      | -       | 7.5000    | 7471B  | PPB   | 28 Jan 2020 12:40:03 | S    | Std  |
| 10.0 - 1        | SD.0                | 60096      | -       | 10.0000   | 7471B  | PPB   | 28 Jan 2020 12:42:19 | S    | Std  |
| ICV90 - 1       | ICV90               | 24189      | 3.9463  | -         | 7471B  | PPB   | 28 Jan 2020 13:08:30 | U    | SMPL |
| ICB90 - 1       | ICB90               | 344        | -0.0135 | -         | 7471B  | PPB   | 28 Jan 2020 13:10:44 | U    | SMPL |
| CCV26 - 1       | CCV26               | 30338      | 4.9674  | -         | 7471B  | PPB   | 28 Jan 2020 13:12:59 | U    | SMPL |
| CCB26 - 1       | CCB26               | 312        | -0.0188 | -         | 7471B  | PPB   | 28 Jan 2020 13:15:14 | U    | SMPL |
| CRA - 1         | CRA                 | 1442       | 0.1689  | -         | 7471B  | PPB   | 28 Jan 2020 13:17:28 | U    | SMPL |
| HighStd - 1     | HighStd             | 61465      | 10.1364 | -         | 7471B  | PPB   | 28 Jan 2020 13:19:43 | U    | SMPL |
| ChkStd - 1      | ChkStd              | 45185      | 7.4329  | -         | 7471B  | PPB   | 28 Jan 2020 13:21:59 | U    | SMPL |
| PB126406BL - 1  | PBS                 | 444        | 0.0031  | -         | 7471B  | PPB   | 28 Jan 2020 13:27:41 | U    | SMPL |
| PB126406BS - 1  | LCSS                | 24769      | 4.0426  | -         | 7471B  | PPB   | 28 Jan 2020 13:29:56 | U    | SMPL |
| L1008-15 - 1    | CL-02-012420        | 5974       | 0.9215  | -         | 7471B  | PPB   | 28 Jan 2020 13:32:12 | U    | SMPL |
| L1008-15DUP - 1 | CL-02-012420DUP     | 5423       | 0.8300  | -         | 7471B  | PPB   | 28 Jan 2020 13:34:28 | U    | SMPL |
| L1008-15MS - 1  | CL-02-012420MS      | 27204      | 4.4470  | -         | 7471B  | PPB   | 28 Jan 2020 13:40:38 | U    | SMPL |
| L1008-15MSD - 1 | CL-02-012420MSD     | 27299      | 4.4627  | -         | 7471B  | PPB   | 28 Jan 2020 13:45:30 | U    | SMPL |
| L1255-01 - 1    | 20200074-COMPOSITE  | 14986      | 2.4180  | -         | 7471B  | PPB   | 28 Jan 2020 13:47:44 | U    | SMPL |
| CCV27 - 1       | CCV27               | 30424      | 4.9817  | -         | 7471B  | PPB   | 28 Jan 2020 13:49:59 | U    | SMPL |
| CCB27 - 1       | CCB27               | 396        | -0.0048 | -         | 7471B  | PPB   | 28 Jan 2020 13:52:14 | U    | SMPL |
| L1258-01 - 1    | SP-1                | 2354       | 0.3203  | -         | 7471B  | PPB   | 28 Jan 2020 13:54:29 | U    | SMPL |
| L1262-01 - 1    | B10(0-15)           | 2660       | 0.3711  | -         | 7471B  | PPB   | 28 Jan 2020 13:56:44 | U    | SMPL |
| L1263-04 - 1    | CTY-SB-0203-0-11    | 30138      | 4.9342  | -         | 7471B  | PPB   | 28 Jan 2020 13:59:00 | U    | SMPL |
| L1268-01 - 1    | 30529               | 350        | -0.0125 | -         | 7471B  | PPB   | 28 Jan 2020 14:01:15 | U    | SMPL |
| L1269-05 - 1    | TP-22-21-COMP       | 1612       | 0.1971  | -         | 7471B  | PPB   | 28 Jan 2020 14:03:31 | U    | SMPL |
| L1269-13 - 1    | TP-25-24-COMP       | 3122       | 0.4478  | -         | 7471B  | PPB   | 28 Jan 2020 14:05:48 | U    | SMPL |
| L1274-01 - 1    | BORING-B1-SAMPLE-S1 | 1153       | 0.1209  | -         | 7471B  | PPB   | 28 Jan 2020 14:08:03 | U    | SMPL |
| L1274-02 - 1    | BORING-B2-SAMPLE-S1 | 696        | 0.0450  | -         | 7471B  | PPB   | 28 Jan 2020 14:10:20 | U    | SMPL |
| L1008-15LX5 - 1 |                     | 1165       | 0.1229  | -         | 7471B  | PPB   | 28 Jan 2020 14:12:37 | U    | SMPL |
| L1016-10 - 1    | SU-04-012320        | 3369       | 0.4889  | -         | 7471B  | PPB   | 28 Jan 2020 14:14:53 | U    | SMPL |
| CCV28 - 1       | CCV28               | 27991      | 4.5777  | -         | 7471B  | PPB   | 28 Jan 2020 14:17:09 | U    | SMPL |
| CCB28 - 1       | CCB28               | 352        | -0.0122 | -         | 7471B  | PPB   | 28 Jan 2020 14:19:24 | U    | SMPL |