



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107385

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/28/2020	12:25
2	CAL2	1	Water	NA	NA	20.2	7.00	01/28/2020	12:26
3	CAL3	1	Water	NA	NA	20.2	10.03	01/28/2020	12:27
4	ICV	1	Water	NA	NA	20.2	6.99	01/28/2020	12:30
5	CCV1	1	Water	NA	NA	20.1	2.01	01/28/2020	12:32
6	L1284-01	1	Water	NA	NA	15.9	5.56	01/28/2020	12:35
7	L1284-01DUP	1	Water	NA	NA	15.8	5.57	01/28/2020	12:36
8	CCV2	1	Water	NA	NA	20.1	12.02	01/28/2020	12:37

VB 107385

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1284 WorkList ID : 135456 Department : Wet-Chemistry Date : 01-28-2020 12:15:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/27/2020	Water	L1284-01-B	pH	Cool 4 deg C	STER12	J21	CELL-POOL-WATER-1ST-Q	01/27/2020	SM 4500-H B

Date/Time 01-28-2020 12:20
Received by: SD
Relinquished by: SD

Date/Time 01-28-2020 13:10
Received by: SD
Relinquished by: SD