



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107388

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/28/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/28/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.02	01/28/2020	15:02
4	ICV	1	Water	NA	NA	20.1	7.00	01/28/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/28/2020	15:06
6	L1285-01	1	Water	NA	NA	13.1	7.01	01/28/2020	15:10
7	L1285-01DUP	1	Water	NA	NA	13.2	7.02	01/28/2020	15:11
8	CCV2	1	Water	NA	NA	20.1	12.02	01/28/2020	15:13

213107388

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1285 WorkList ID : 135466 Department : Wet-Chemistry Date : 01-28-2020 14:40:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2020	Water	L1285-01-E	pH	Cool 4 deg C	TULL02	J12	MH-725	01/28/2020	9040C

Date/Time 01-28-2020 14:45
Received by: Jo
Relinquished by: Ed

Date/Time 01-28-2020 15:30
Received by: [Signature]
Relinquished by: Jo