

245.1


LB107389

Instrument ID : CV1

Linear

 μ Abs.:

37029



A= 0.0000e+000

B= 2.6589e-004

C= -4.5108e-002

Rho= 0.9994301

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	168	0.000	168				
0.05	0.050	0.044	-0.006	336	0.0 %	336				
0.2	0.200	0.188	-0.012	875	0.0 %	875				
2.5	2.500	2.437	-0.063	9335	0.0 %	9335				
5.0	5.000	5.025	0.025	19070	0.0 %	19070				
7.5	7.500	7.756	0.256	29338	0.0 %	29338				
10.0	10.000	9.801	-0.199	37029	0.0 %	37029				

LB107389 INSTRUMENT ID:CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Jan 2020 15:16:16

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50.0	168	-	0.0000245.1	PPB	28 Jan 2020 16:41:56	S	Std
0.05 - 1	50.05	336	-	0.0500245.1	PPB	28 Jan 2020 16:44:13	S	Std
0.2 - 1	50.2	875	-	0.2000245.1	PPB	28 Jan 2020 16:46:30	S	Std
2.5 - 1	52.5	9335	-	2.5000245.1	PPB	28 Jan 2020 16:48:47	S	Std
5.0 - 1	55.0	19070	-	5.0000245.1	PPB	28 Jan 2020 16:51:03	S	Std
7.5 - 1	57.5	29338	-	7.5000245.1	PPB	28 Jan 2020 16:53:19	S	Std
10.0 - 1	510.0	37029	-	10.0000245.1	PPB	28 Jan 2020 16:58:45	S	Std
ICV91 - 1	ICV91	15772	4.1485	-245.1	PPB	28 Jan 2020 17:02:08	U	SMPL
ICB91 - 1	ICB91	163	-0.0018	-245.1	PPB	28 Jan 2020 17:04:23	U	SMPL
CCV29 - 1	CCV29	19757	5.2081	-245.1	PPB	28 Jan 2020 17:06:38	U	SMPL
CCB29 - 1	CCB29	159	-0.0028	-245.1	PPB	28 Jan 2020 17:08:53	U	SMPL
CRA - 1	CRA	338	0.0448	-245.1	PPB	28 Jan 2020 17:11:07	U	SMPL
HighStd - 1	HighStd	36850	9.7530	-245.1	PPB	28 Jan 2020 17:13:22	U	SMPL
ChkStd - 1	ChkStd	27367	7.2316	-245.1	PPB	28 Jan 2020 17:15:37	U	SMPL
PB126413BL - 1	PBW	142	-0.0074	-245.1	PPB	28 Jan 2020 17:17:53	U	SMPL
PB126413BS - 1	LCSW	15855	4.1706	-245.1	PPB	28 Jan 2020 17:20:08	U	SMPL
L1273-01 - 1	1-BSMT-POINT-SOURCE	19	-0.0401	-245.1	PPB	28 Jan 2020 17:33:26	U	SMPL
L1273-01DUP - 1	1-BSMT-POINT-SOURCEDUP	23	-0.0390	-245.1	PPB	28 Jan 2020 17:35:41	U	SMPL
L1273-01MS - 1	1-BSMT-POINT-SOURCEMS	15998	4.2086	-245.1	PPB	28 Jan 2020 17:46:04	U	SMPL
L1273-01MSD - 1	1-BSMT-POINT-SOURCEMSD	16221	4.2679	-245.1	PPB	28 Jan 2020 17:48:20	U	SMPL
L1273-02 - 1	1F-BSMT-POINT-SOURCE	1	-0.0448	-245.1	PPB	28 Jan 2020 17:50:35	U	SMPL
CCV30 - 1	CCV30	19739	5.2033	-245.1	PPB	28 Jan 2020 17:52:50	U	SMPL
CCB30 - 1	CCB30	158	-0.0031	-245.1	PPB	28 Jan 2020 17:55:05	U	SMPL
L1273-03 - 1	2-1ST-FL-PREK-CR-SINK	10	-0.0424	-245.1	PPB	28 Jan 2020 17:57:20	U	SMPL
L1273-04 - 1	2F-1ST-FL-PREK-CR-SINK	11	-0.0422	-245.1	PPB	28 Jan 2020 17:59:35	U	SMPL
L1273-05 - 1	3-1ST-FL-PREK-BR-SINK	22	-0.0393	-245.1	PPB	28 Jan 2020 18:01:50	U	SMPL
L1273-06 - 1	3F-1ST-FL-PREK-BR-SINK	13	-0.0417	-245.1	PPB	28 Jan 2020 18:04:06	U	SMPL
L1273-07 - 1	4-4TH-FL-HALLWAY-FTN	10	-0.0424	-245.1	PPB	28 Jan 2020 18:06:22	U	SMPL
L1273-08 - 1	4F-4TH-FL-HALLWAY-FTN	17	-0.0406	-245.1	PPB	28 Jan 2020 18:08:38	U	SMPL
L1273-09 - 1	5-4TH-FL-MS-GREEN-AREA-FTN	7	-0.0432	-245.1	PPB	28 Jan 2020 18:10:54	U	SMPL
L1273-10 - 1	5F-4TH-FL-MS-GREEN-AREA-FTN	11	-0.0422	-245.1	PPB	28 Jan 2020 18:13:11	U	SMPL
L1273-11 - 1	6-4TH-FL-MS-RED-AREA-FTN	13	-0.0417	-245.1	PPB	28 Jan 2020 18:15:28	U	SMPL
L1273-12 - 1	6F-4TH-FL-MS-RED-AREA-FTN	14	-0.0414	-245.1	PPB	28 Jan 2020 18:17:44	U	SMPL
CCV31 - 1	CCV31	19828	5.2270	-245.1	PPB	28 Jan 2020 18:19:59	U	SMPL
CCB31 - 1	CCB31	158	-0.0031	-245.1	PPB	28 Jan 2020 18:22:14	U	SMPL
L1273-13 - 1	7-4TH-FL-MS-BLUE-AREA-FTN	4	-0.0440	-245.1	PPB	28 Jan 2020 18:24:29	U	SMPL
L1273-14 - 1	7F-4TH-FL-MS-BLUE-AREA-FTN	15	-0.0411	-245.1	PPB	28 Jan 2020 18:26:45	U	SMPL
L1273-15 - 1	8-RM-485-TEACHERS-SINK	30	-0.0371	-245.1	PPB	28 Jan 2020 18:29:00	U	SMPL
L1273-16 - 1	8F-RM-485-TEACHERS-SINK	1	-0.0448	-245.1	PPB	28 Jan 2020 18:31:16	U	SMPL
L1273-17 - 1	9-RM-481-SINK-1	19	-0.0401	-245.1	PPB	28 Jan 2020 18:33:31	U	SMPL
L1273-18 - 1	9F-RM-481-SINK-1	26	-0.0382	-245.1	PPB	28 Jan 2020 18:35:47	U	SMPL
L1273-19 - 1	10-RM-481-SINK-2	18	-0.0403	-245.1	PPB	28 Jan 2020 18:38:04	U	SMPL
L1273-20 - 1	10F-RM-481-SINK-2	23	-0.0390	-245.1	PPB	28 Jan 2020 18:40:21	U	SMPL
L1273-21 - 1	11-RM-480-SINK-2	15	-0.0411	-245.1	PPB	28 Jan 2020 18:42:37	U	SMPL
L1273-22 - 1	11F-RM-480-SINK-2	11	-0.0422	-245.1	PPB	28 Jan 2020 18:44:53	U	SMPL
CCV32 - 1	CCV32	19484	5.1355	-245.1	PPB	28 Jan 2020 18:47:08	U	SMPL
CCB32 - 1	CCB32	161	-0.0023	-245.1	PPB	28 Jan 2020 18:49:24	U	SMPL
PB126414BL - 1	PBW	183	0.0035	-245.1	PPB	28 Jan 2020 18:51:39	U	SMPL
PB126414BS - 1	LCSW	15718	4.1342	-245.1	PPB	28 Jan 2020 18:53:54	U	SMPL
L1256-01 - 1	TANK-1	67	-0.0273	-245.1	PPB	28 Jan 2020 18:56:10	U	SMPL
L1256-01DUP - 1	TANK-1DUP	99	-0.0188	-245.1	PPB	28 Jan 2020 18:58:25	U	SMPL
L1256-01MS - 1	TANK-1MS	14426	3.7906	-245.1	PPB	28 Jan 2020 19:07:56	U	SMPL
L1256-01MSD - 1	TANK-1MSD	15315	4.0270	-245.1	PPB	28 Jan 2020 19:12:59	U	SMPL
L1272-01 - 1	TI-658	293	0.0328	-245.1	PPB	28 Jan 2020 19:15:15	U	SMPL
L1273-01LX5 - 1		0	-0.0451	-245.1	PPB	28 Jan 2020 19:17:31	U	SMPL
L1256-01LX5 - 1		32	-0.0366	-245.1	PPB	28 Jan 2020 19:19:49	U	SMPL
CCV33 - 1	CCV33	19592	5.1642	-245.1	PPB	28 Jan 2020 19:22:04	U	SMPL
CCB33 - 1	CCB33	171	0.0004	-245.1	PPB	28 Jan 2020 19:24:20	U	SMPL