



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107391

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/28/2020	12:40
2	L1265-01		1	9	100.00	>212.0	>212.0	01/28/2020	13:05
3	L1265-01DUP		1	9	100.00	>212.0	>212.0	01/28/2020	13:35
4	L1266-01		1	8	100.00	>212.0	>212.0	01/28/2020	14:10
5	L1267-03		1	7	100.00	>212.0	>212.0	01/28/2020	14:45
6	L1267-09		1	8	100.00	>212.0	>212.0	01/28/2020	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB107391

WorkList Name : FP12820 WorkList ID : 135449 Department : Wet-Chemistry Date : 01-28-2020 10:45:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/06/2020	Water	L1265-01	Flash Point	Cool 4 deg C	PARS15	I63	CTY-WC-L-01	01/23/2020	1010A
01/31/2020	Water	L1266-01	Flash Point	Cool 4 deg C	PSEG01	I41	40652	01/24/2020	1010A
01/31/2020	Water	L1267-03	Flash Point	Cool 4 deg C	PSEG01	I41	40329	01/24/2020	1010A
01/31/2020	Water	L1267-09	Flash Point	Cool 4 deg C	PSEG01	I41	COMP-40328-505-670	01/24/2020	1010A

Date/Time 01/28/20 12:15
 Received by: KS
 Relinquished by: JP

Date/Time 01/28/20 15:25
 Received by: JP
 Relinquished by: KS