



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107392

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/28/2020	11:20
2	L1268-06		1	8	70.00	158.0	158.0	01/28/2020	11:45
3	L1268-06DUP		1	8	68.00	154.4	154.4	01/28/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 107392

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP12820-1 WorkList ID : 135450 Department : Wet-Chemistry Date : 01-28-2020 10:46:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2020	Solid	L1268-06	A	Cool 4 deg C	PSEG01	L41	17562	01/24/2020	1010A

Date/Time 01/28/20 10:55
Received by: KS
Relinquished by: KS

Date/Time 01/28/20 12:15
Received by: KS
Relinquished by: KS