

LB161393

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,

Mountainside, NJ 07092

Reviewed by : AK

Instrument ID : Konelab

1/28/2020 15:32

Test: Total CN

ugil

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.875	0.0	0.090	
ICB1	-0.096	0.0	0.002	
CCV1	244.364	0.0	0.229	
CCB1	-0.212	0.0	0.002	
PB126410BL	-0.403	0.0	0.002	
PB126410BS	96.250	0.0	0.092	
MIDPB106410	246.327	0.0	0.231	
L1161-10	2.824	0.0	0.005	
L1161-11	3.293	0.0	0.006	
CCV2	256.673	0.0	0.240	
CCB2	-0.124	0.0	0.002	

AK
01-28-20

N	11
Mean	85.797
SD	111.1682
CV%	129.57

AK

Aquakem v. 7.2AQ1

Results from time period:

Tue Jan 28 15:24:26 2020

Tue Jan 28 15:27:50 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.5175	µg/l	1/28/2020 8:51:08	
5.OPPBCN	A	Total CN	P	4.4999	µg/l	1/28/2020 8:51:09	
10PPBCN	A	Total CN	P	9.6385	µg/l	1/28/2020 8:51:10	
50PPBCN	A	Total CN	P	49.4499	µg/l	1/28/2020 8:51:11	
100PPBCN	A	Total CN	P	99.7258	µg/l	1/28/2020 8:51:12	
250PPBCN	A	Total CN	P	254.1627	µg/l	1/28/2020 8:51:13	
500PPBCN	A	Total CN	P	498.0407	µg/l	1/28/2020 8:51:14	
ICV1	S	Total CN	P	94.8753	µg/l	1/28/2020 15:24:26	
ICB1	S	Total CN	P	-0.0963	µg/l	1/28/2020 15:24:27	
CCV1	S	Total CN	P	244.3643	µg/l	1/28/2020 15:24:28	
CCB1	S	Total CN	P	-0.2122	µg/l	1/28/2020 15:24:29	
PB126410BL	S	Total CN	P	-0.4032	µg/l	1/28/2020 15:24:30	
PB126410BS	S	Total CN	P	96.2505	µg/l	1/28/2020 15:24:31	
MIDPB106410	S	Total CN	P	246.3274	µg/l	1/28/2020 15:24:32	
L1161-10	S	Total CN	P	2.8236	µg/l	1/28/2020 15:24:35	
L1161-11	S	Total CN	P	3.293	µg/l	1/28/2020 15:24:36	
CCV2	S	Total CN	P	256.6734	µg/l	1/28/2020 15:27:49	
CCB2	S	Total CN	P	-0.1235	µg/l	1/28/2020 15:27:50	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AL

Instrument ID : Konelab

1/28/2020 15:40


Test Total CN

Accepted 1/28/2020 8:54

Factor 1079
Bias 0.002

Coeff. of det. 0.999892

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.5175	0.0000	
2	5.0PPBCN	0.007	4.4999	5.0000	
3	10PPBCN	0.011	9.6385	10.0000	
4	50PPBCN	0.048	49.4499	50.0000	
5	100PPBCN	0.095	99.7258	100.0000	
6	250PPBCN	0.238	254.1627	250.0000	
7	500PPBCN	0.464	498.0407	500.0000	