



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107394

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/28/2020	17:35
2	CAL2	1	Water	NA	NA	20.1	7.01	01/28/2020	17:36
3	CAL3	1	Water	NA	NA	20.1	10.03	01/28/2020	17:38
4	ICV	1	Water	NA	NA	20.1	7.00	01/28/2020	17:40
5	CCV1	1	Water	NA	NA	20.2	2.01	01/28/2020	17:41
6	L1013-12	1	Solid	20.02	20	20.5	7.33	01/28/2020	17:42
7	L1287-05	1	Solid	20.03	20	21.5	7.88	01/28/2020	17:43
8	L1287-06	1	Solid	20.04	20	21.4	7.96	01/28/2020	17:45
9	L1289-06	1	Solid	20.02	20	20.2	5.84	01/28/2020	17:46
10	L1289-14	1	Solid	20.03	20	20.1	5.85	01/28/2020	17:48
11	L1289-22	1	Solid	20.02	20	19.7	5.83	01/28/2020	17:50
12	L1289-30	1	Solid	20.03	20	20.1	5.75	01/28/2020	18:00
13	L1289-38	1	Solid	20.02	20	19.9	5.18	01/28/2020	18:02
14	L1289-38DUP	1	Solid	20.03	20	19.8	5.19	01/28/2020	18:03
15	CCV2	1	Water	NA	NA	20.1	12.01	01/28/2020	18:05

VB107394

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1287 WorkList ID : 135460 Department : Wet-Chemistry Date : 01-28-2020 13:23:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2020	Solid	L1013-12	Corrosivity	Cool 4 deg C	PSEG05		PL-05-012820	01/28/2020	9045D
01/28/2020	Solid	L1287-05	Corrosivity	Cool 4 deg C	TULL02	J13	AOC-3-EXC-PIT-(0-5)-A	01/28/2020	9045D
01/28/2020	Solid	L1287-06	Corrosivity	Cool 4 deg C	TULL02	J13	AOC-3-EXC-PIT-(0-5)-B	01/28/2020	9045D
01/27/2020	Solid	L1289-06	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-26-27-COMP	01/27/2020	9045D
01/27/2020	Solid	L1289-14	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-29-30-COMP	01/27/2020	9045D
01/27/2020	Solid	L1289-22	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-32-33-COMP	01/27/2020	9045D
01/28/2020	Solid	L1289-30	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-35-34-COMP	01/28/2020	9045D
01/28/2020	Solid	L1289-38	Corrosivity	Cool 4 deg C	PSEG01	I42	TP-3-COMP	01/28/2020	9045D

Date/Time 01.28.2020 16:10
 Received by: Jg
 Relinquished by: JR

Date/Time 01.28.2020 18:40
 Received by: JR
 Relinquished by: Jg