



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107403

Slope : 98.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/29/2020	11:35
2	CAL2	1	Water	NA	NA	20.2	7.00	01/29/2020	11:36
3	CAL3	1	Water	NA	NA	20.2	10.03	01/29/2020	11:37
4	ICV	1	Water	NA	NA	20.2	7.01	01/29/2020	11:40
5	CCV1	1	Water	NA	NA	20.1	2.01	01/29/2020	11:42
6	L1291-01	1	Water	NA	NA	20.3	7.16	01/29/2020	11:43
7	L1291-01DUP	1	Water	NA	NA	20.4	7.17	01/29/2020	11:44
8	CCV2	1	Water	NA	NA	20.2	12.01	01/29/2020	11:45

107403

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1291 WorkList ID : 135493 Department : Wet-Chemistry Date : 01-29-2020 11:29:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2020	Water	L1291-01	pH	Cool 4 deg C	TULL02	J31	MH-725-(SETTLED)	01/29/2020	9040C

Date/Time 01-29-2020 11:30
Received by: [Signature]
Relinquished by: _____

Date/Time 01-29-2020 12:30
Received by: [Signature]
Relinquished by: Jo