



# PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/30/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:40  
In Date: 01/29/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:42  
Out Date: 01/30/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB107407

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|-------------------------|---------------------------|---------|----------|
| L1300-01 | C5AK0           | 1      | 1.15           | 9.66                    | 7.82                      | 78.4    |          |
| L1300-02 | C5AK2           | 2      | 1.12           | 9.72                    | 7.96                      | 79.5    |          |
| L1300-03 | C5AK4           | 3      | 1.14           | 9.9                     | 7.74                      | 75.3    |          |
| L1300-04 | C5AN4           | 4      | 1.18           | 9.92                    | 8.1                       | 79.2    |          |
| L1300-05 | C5AN6           | 5      | 1.16           | 9.93                    | 7.91                      | 77.0    |          |
| L1300-06 | C5AN8           | 6      | 1.12           | 9.68                    | 7.59                      | 75.6    |          |
| L1300-07 | C5AP0           | 7      | 1.14           | 9.6                     | 7.81                      | 78.8    |          |
| L1300-08 | C5AP2           | 8      | 1.13           | 9.94                    | 8.14                      | 79.6    |          |
| L1300-09 | C5AP4           | 9      | 1.18           | 9.72                    | 7.96                      | 79.4    |          |
| L1300-10 | C5AP6           | 10     | 1.17           | 9.86                    | 8.08                      | 79.5    |          |
| L1300-11 | C5AR0           | 11     | 1.14           | 9.95                    | 8.01                      | 78.0    |          |
| L1300-12 | C5AR2           | 12     | 1.14           | 9.86                    | 7.87                      | 77.2    |          |
| L1300-13 | C5AR4           | 13     | 1.18           | 9.68                    | 7.39                      | 73.1    |          |
| L1300-14 | C5AR6           | 14     | 1.13           | 9.77                    | 7.97                      | 79.2    |          |
| L1300-15 | C5AR6MS         | 15     | 1.13           | 9.77                    | 7.97                      | 79.2    |          |
| L1300-16 | C5AR6MSD        | 16     | 1.13           | 9.77                    | 7.97                      | 79.2    |          |
| L1300-17 | C5AR8           | 17     | 1.18           | 9.7                     | 7.44                      | 73.5    |          |
| L1300-18 | C5AS0           | 18     | 1.19           | 9.59                    | 7.8                       | 78.7    |          |
| L1300-19 | C5AS8           | 19     | 1.14           | 9.9                     | 7.67                      | 74.5    |          |
| L1300-20 | C5AT0           | 20     | 1.18           | 9.98                    | 8.12                      | 78.9    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$