



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/30/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:13
In Date: 01/29/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 01/30/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107418

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1004-16	BU-01-012920	1	1.14	9.57	8.9	92.1	
L1004-17	BU-01-012920	2	1.15	9.79	9.18	92.9	
L1014-11	OR-03-012920	3	1.14	9.78	9.1	92.1	
L1014-12	OR-03-012920	4	1.13	9.91	9.47	95.0	
L1014-13	OR-03-012920-1	5	1.13	9.82	9.32	94.2	
L1014-14	OR-03-012920-2	6	1.14	9.56	9.17	95.4	
L1296-01	RBR-51	7	1.16	9.7	8.95	91.2	
L1297-01	PCB-0122220-01	8	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1297-02	PCB-0122220-02	9	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1297-03	PCB-0122220-03	10	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1297-04	PCB-012220-04	11	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L1298-01	RBR-54	12	1.13	9.63	8.58	87.6	
L1307-01	SP-NORTH-SIDE	13	1.15	9.78	8.16	81.2	
L1307-02	SP-NORTH-SIDE	14	1.14	9.58	7.98	81.0	
L1307-03	SP-SOUTH-SIDE	15	1.14	9.71	8.19	82.3	
L1307-04	SP-SOUTH-SIDE	16	1.13	9.77	8.4	84.1	
L1308-01	STOCK-PILE	17	1.15	9.93	8.26	81.0	
L1309-01	CTY-WC-S-SB-0102	18	1.15	9.63	7.5	74.9	
L1310-01	CTY-SB-02-79-80	19	1.13	9.82	8.33	82.9	
L1310-02	CTY-SB-0102-0-80	20	1.14	9.85	7.19	69.5	
L1310-03	CTY-SB-10102-0-80	21	1.12	9.85	7.32	71.0	
L1313-01	RBR-36	22	1.14	9.67	8.19	82.6	
L1313-02	124018-A	23	1.14	9.97	7.61	73.3	
L1313-03	RBR-56	24	1.13	9.62	8.06	81.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$