



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107419

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/29/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	01/29/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/29/2020	17:19
4	ICV	1	Water	NA	NA	20.1	7.00	01/29/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	01/29/2020	17:22
6	L1004-18	1	Solid	20.03	20	21.7	6.72	01/29/2020	17:25
7	L1014-15	1	Solid	20.02	20	20.8	8.26	01/29/2020	17:26
8	L1309-01	1	Solid	20.03	20	20.4	9.83	01/29/2020	17:30
9	L1309-01DUP	1	Solid	20.04	20	20.5	9.84	01/29/2020	17:32
10	CCV2	1	Water	NA	NA	20.1	12.02	01/29/2020	17:35

107419

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1309 WorkList ID : 135509 Department : Wet-Chemistry Date : 01-29-2020 14:30:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2020	Solid	L1004-18	Corrosivity	Cool 4 deg C	PSEG05		BU-01-012920	01/29/2020	9045D
01/29/2020	Solid	L1014-15	Corrosivity	Cool 4 deg C	PSEG05		OR-03-012920	01/29/2020	9045D
01/20/2020	Solid	L1309-01	Corrosivity	Cool 4 deg C	PARS15	J41	CTY-WC-S-SB-0102	01/20/2020	9045D

Date/Time 01-29-2020 16:10
 Received by: JP
 Relinquished by: JP

Date/Time 01-29-2020 18:20
 Received by: JP
 Relinquished by: JP