

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 01/31/2020

OVEN IN TEMP1: 98 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 01/30/2020 16:00

DATE & TIME OUT1: 01/31/2020 07:10

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 01/31/2020 08:00

DATE & TIME OUT2: 01/31/2020 09:35

Run Number: LB107431

BalanceID: EC DC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB107431BL	LB107431BL	79.5867	79.5867	100	79.5867	79.5867	79.5867	0.0000	0
2	L1287-11	AOC-3-EXC-PIT- (0-5) -A	79.5048	79.5048	100	79.5170	79.5170	79.5170	0.0122	122
3	L1287-11DUP	AOC-3-EXC-PIT- (0-5) -ADUP	78.4793	78.4793	100	78.4917	78.4917	78.4917	0.0124	124
4	L1287-12	AOC-3-EXC-PIT- (0-5) -B	76.7384	76.7384	100	76.7497	76.7497	76.7497	0.0113	113

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

HB107431

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS L1287

WorkList ID : 135542

Department : Wet-Chemistry

Date : 01-30-2020 11:25:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
02/04/2020	Solid	L1287-11	ASTM TS	Cool 4 deg C	TULL02	J13	AOC-3-EXC-PIT-(0-5)-A	01/28/2020	SM2540 B
02/04/2020	Solid	L1287-12	ASTM TS	Cool 4 deg C	TULL02	J13	AOC-3-EXC-PIT-(0-5)-B	01/28/2020	SM2540 B

Date/Time 01-30-2020 11:30
Received by: J0
Relinquished by: J0

Date/Time 01-30-2020 16:11:10
Received by: J0
Relinquished by: J0