



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107434

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/29/2020	09:45
2	L1283-01		1	8	100.00	>212.0	>212.0	01/29/2020	10:20
3	L1283-01DUP		1	9	100.00	>212.0	>212.0	01/30/2020	10:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12920 WorkList ID : 135488 Department : Wet-Chemistry Date : 01-29-2020 09:14:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/31/2020	Water	L1283-01	Flash Point	Cool 4 deg C	PSEG01	I42	VB16209	01/28/2020	1010A

Date/Time 01/29/20 09:20
 Received by: KS
 Relinquished by: DK

Date/Time 01/29/20 10:55
 Received by: DK
 Relinquished by: KS