



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB107435

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP81452
COD calibration std. 100 ppm	WP81448
COD calibration std. 10 ppm	WP81446
COD calibration std. 150 ppm	WP81450
COD calibration std. 50 ppm	WP81447
COD calibration std. 0 ppm	WP81445
COD Digestion Vials Low Level 0-150Mg/L	W2621
COD ICV-LCS std, 50ppm	WP81658
COD CCV std, 50ppm	WP81657
COD LOD std, 5ppm	WP81627
COD Digestion Vials Low Level 0-150Mg/L	W2622

Temp In(C): 148	Date In: 01/30/2020	Time In: 09:30
Temp Out(C): 148	Date Out: 01/30/2020	Time Out: 11:30

Intercept: -0.4824

Slope: 1.0142

Regression: 0.9998

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	01/24/2020	15:10
2	CAL2	10	1	Water	10.000	01/24/2020	15:10
3	CAL3	50	1	Water	50.000	01/24/2020	15:11
4	CAL4	100	1	Water	99.000	01/24/2020	15:11
5	CAL5	150	1	Water	153.000	01/24/2020	15:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB107435

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	48.000	47.804	01/24/2020	15:12
2	ICB				1	Water	0.000	0.476	01/24/2020	15:13
3	CCV1	50			1	Water	48.000	47.804	01/30/2020	12:10
4	CCB1				1	Water	0.000	0.476	01/30/2020	12:10
5	LB107435BL				1	Water	0.000	0.476	01/30/2020	12:11
6	LB107435BS	50			1	Water	51.000	50.762	01/30/2020	12:11
7	L1161-09				1	Water	3.000	3.434	01/30/2020	12:12
8	L1299-01				1	Water	0.000	0.476	01/30/2020	12:12
9	L1299-01DUP				1	Water	0.000	0.476	01/30/2020	12:13
10	L1299-01MS	50			1	Water	50.000	49.776	01/30/2020	12:13
11	L1299-01MSD	50			1	Water	49.000	48.790	01/30/2020	12:14
12	CCV2	50			1	Water	48.000	47.804	01/30/2020	12:14
13	CCB2				1	Water	0.000	0.476	01/30/2020	12:15