

LB107449


7470A

Instrument ID = CV1

Linear

μ Abs.:

40656



A= 0.0000e+000

B= 2.4406e-004

C= -1.3480e-002

Rho= 0.9991671

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.042	0.042	228	0.000	228				
0.2	0.200	0.233	0.033	1010	0.0 %	1010				
2.5	2.500	2.488	-0.012	10250	0.0 %	10250				
5.0	5.000	4.764	-0.236	19576	0.0 %	19576				
7.5	7.500	7.763	0.263	31863	0.0 %	31863				
10.0	10.000	9.909	-0.091	40656	0.0 %	40656				

LB107444 INSTRUMENT ID:CV1

Method: 7470A		Operator: Admin		Date of Analysis: 30 Jan 2020 14:50:30					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	228	-	0.0000	7470A	PPB	30 Jan 2020 14:50:51	S	Std
0.2 - 1	SD-2	1010	-	0.2000	7470A	PPB	30 Jan 2020 14:52:58	S	Std
2.5 - 1	SD-5	10250	-	2.5000	7470A	PPB	30 Jan 2020 14:55:05	S	Std
5.0 - 1	SD-10	19576	-	5.0000	7470A	PPB	30 Jan 2020 14:57:13	S	Std
7.5 - 1	SD-15	31863	-	7.5000	7470A	PPB	30 Jan 2020 14:59:20	S	Std
10.0 - 1	SD-20	40656	-	10.0000	7470A	PPB	30 Jan 2020 15:01:26	S	Std
ICV94 - 1	ICV94	16938	4.1205	-	7470A	PPB	30 Jan 2020 15:07:22	U	SMPL
ICB94 - 1	ICB94	138	0.0202	-	7470A	PPB	30 Jan 2020 15:09:28	U	SMPL
CCV46 - 1	CCV46	20909	5.0897	-	7470A	PPB	30 Jan 2020 15:11:34	U	SMPL
CCB46 - 1	CCB46	126	0.0173	-	7470A	PPB	30 Jan 2020 15:13:39	U	SMPL
CRA - 1	CRA	1012	0.2335	-	7470A	PPB	30 Jan 2020 15:15:45	U	SMPL
HighStd - 1	HighStd	41614	10.1430	-	7470A	PPB	30 Jan 2020 15:17:50	U	SMPL
ChkStd - 1	ChkStd	28578	6.9614	-	7470A	PPB	30 Jan 2020 15:22:57	U	SMPL
PB126464BL - 1	PBW	182	0.0309	-	7470A	PPB	30 Jan 2020 15:25:03	U	SMPL
PB126464BS - 1	LCSW	16610	4.0404	-	7470A	PPB	30 Jan 2020 15:27:10	U	SMPL
L1236-15 - 1	40218-W	168	0.0275	-	7470A	PPB	30 Jan 2020 15:29:17	U	SMPL
L1236-15DUP - 1	40218-WDUP	144	0.0217	-	7470A	PPB	30 Jan 2020 15:31:24	U	SMPL
L1236-15MS - 1	40218-WMS	16246	3.9516	-	7470A	PPB	30 Jan 2020 15:33:31	U	SMPL
L1236-15MSD - 1	40218-WMSD	16417	3.9933	-	7470A	PPB	30 Jan 2020 15:35:39	U	SMPL
PB126465BL - 1	PBW	144	0.0217	-	7470A	PPB	30 Jan 2020 15:37:45	U	SMPL
CCV47 - 1	CCV47	20183	4.9125	-	7470A	PPB	30 Jan 2020 15:39:51	U	SMPL
CCB47 - 1	CCB47	147	0.0224	-	7470A	PPB	30 Jan 2020 15:41:56	U	SMPL
PB126465BS - 1	LCSW	18016	4.3836	-	7470A	PPB	30 Jan 2020 15:46:52	U	SMPL
L1004-18 - 1	BU-01-012920	144	0.0217	-	7470A	PPB	30 Jan 2020 15:48:58	U	SMPL
L1004-18DUP - 1	BU-01-012920DUP	160	0.0256	-	7470A	PPB	30 Jan 2020 15:51:03	U	SMPL
L1004-18MS - 1	BU-01-012920MS	16217	3.9445	-	7470A	PPB	30 Jan 2020 15:53:10	U	SMPL
L1004-18MSD - 1	BU-01-012920MSD	16547	4.0250	-	7470A	PPB	30 Jan 2020 15:55:16	U	SMPL
L1014-15 - 1	OR-03-012920	145	0.0219	-	7470A	PPB	30 Jan 2020 15:57:23	U	SMPL
L1295-01 - 1	FM-406	137	0.0200	-	7470A	PPB	30 Jan 2020 15:59:30	U	SMPL
L1295-02 - 1	FM-1296	147	0.0224	-	7470A	PPB	30 Jan 2020 16:01:37	U	SMPL
L1295-03 - 1	N48838	129	0.0180	-	7470A	PPB	30 Jan 2020 16:03:45	U	SMPL
L1309-01 - 1	CTY-WC-S-SB-0102	160	0.0256	-	7470A	PPB	30 Jan 2020 16:05:51	U	SMPL
CCV48 - 1	CCV48	20236	4.9254	-	7470A	PPB	30 Jan 2020 16:07:57	U	SMPL
CCB48 - 1	CCB48	153	0.0239	-	7470A	PPB	30 Jan 2020 16:10:03	U	SMPL
PB126481BL - 1	PBW	154	0.0241	-	7470A	PPB	30 Jan 2020 16:12:09	U	SMPL
PB126481BS - 1	LCSW	13640	3.3156	-	7470A	PPB	30 Jan 2020 16:14:15	U	SMPL
L1283-01 - 1	VB16209	165	0.0268	-	7470A	PPB	30 Jan 2020 16:16:21	U	SMPL
L1283-01DUP - 1	VB16209DUP	165	0.0268	-	7470A	PPB	30 Jan 2020 16:18:27	U	SMPL
L1283-01MS - 1	VB16209MS	15630	3.8012	-	7470A	PPB	30 Jan 2020 16:20:34	U	SMPL
L1283-01MSD - 1	VB16209MSD	15533	3.7776	-	7470A	PPB	30 Jan 2020 16:22:41	U	SMPL
L1294-07 - 1	M1	80	0.0060	-	7470A	PPB	30 Jan 2020 16:24:48	U	SMPL
L1304-01 - 1	0075-COMP-ELUTIRATE	134	0.0192	-	7470A	PPB	30 Jan 2020 16:26:55	U	SMPL
L1304-04 - 1	0075-COMP-ELUTIRATE	103	0.0117	-	7470A	PPB	30 Jan 2020 16:29:02	U	SMPL
PB126450BL - 1	PBW	106	0.0124	-	7470A	PPB	30 Jan 2020 16:31:09	U	SMPL
CCV49 - 1	CCV49	19952	4.8561	-	7470A	PPB	30 Jan 2020 16:33:15	U	SMPL
CCB49 - 1	CCB49	151	0.0234	-	7470A	PPB	30 Jan 2020 16:35:21	U	SMPL
PB126450BS - 1	LCSW	16191	3.9382	-	7470A	PPB	30 Jan 2020 16:37:27	U	SMPL
L1013-12 - 1	PL-05-012820	126	0.0173	-	7470A	PPB	30 Jan 2020 16:39:33	U	SMPL
L1013-12DUP - 1	PL-05-012820DUP	116	0.0148	-	7470A	PPB	30 Jan 2020 16:41:39	U	SMPL
L1013-12MS - 1	PL-05-012820MS	16693	4.0607	-	7470A	PPB	30 Jan 2020 16:43:46	U	SMPL
L1013-12MSD - 1	PL-05-012820MSD	16595	4.0368	-	7470A	PPB	30 Jan 2020 16:45:52	U	SMPL
L1287-05 - 1	AOC-3-EXC-PIT-(0-5)-A	136	0.0197	-	7470A	PPB	30 Jan 2020 16:47:59	U	SMPL
L1287-06 - 1	AOC-3-EXC-PIT-(0-5)-B	118	0.0153	-	7470A	PPB	30 Jan 2020 16:50:06	U	SMPL
L1289-06 - 1	TP-26-27-COMP	118	0.0153	-	7470A	PPB	30 Jan 2020 16:52:14	U	SMPL
L1289-14 - 1	TP-29-30-COMP	117	0.0151	-	7470A	PPB	30 Jan 2020 16:54:21	U	SMPL
L1289-22 - 1	TP-32-33-COMP	119	0.0156	-	7470A	PPB	30 Jan 2020 16:56:28	U	SMPL
CCV50 - 1	CCV50	18726	4.5569	-	7470A	PPB	30 Jan 2020 16:58:34	U	SMPL
CCB50 - 1	CCB50	109	0.0131	-	7470A	PPB	30 Jan 2020 17:00:40	U	SMPL
L1289-30 - 1	TP-35-34-COMP	113	0.0141	-	7470A	PPB	30 Jan 2020 17:02:46	U	SMPL
L1289-38 - 1	TP-3-COMP	129	0.0180	-	7470A	PPB	30 Jan 2020 17:04:52	U	SMPL
L1236-15LX5 - 1		27	-0.0069	-	7470A	PPB	30 Jan 2020 17:06:58	U	SMPL
L1013-12LX5 - 1		22	-0.0081	-	7470A	PPB	30 Jan 2020 17:09:04	U	SMPL
L1283-01LX5 - 1		38	-0.0042	-	7470A	PPB	30 Jan 2020 17:11:10	U	SMPL
L1004-18LX5 - 1		56	0.0002	-	7470A	PPB	30 Jan 2020 17:13:17	U	SMPL
PB126465TB - 1	PB126446TB	125	0.0170	-	7470A	PPB	30 Jan 2020 17:15:24	U	SMPL
PB126464TB - 1	PB126373TB	157	0.0248	-	7470A	PPB	30 Jan 2020 17:17:32	U	SMPL
PB126450TB - 1	PB126412TB	142	0.0212	-	7470A	PPB	30 Jan 2020 17:19:40	U	SMPL
CCV51 - 1	CCV51	19116	4.6520	-	7470A	PPB	30 Jan 2020 17:21:46	U	SMPL
CCB51 - 1	CCB51	151	0.0234	-	7470A	PPB	30 Jan 2020 17:23:53	U	SMPL