



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107451

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/30/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/30/2020	17:32
3	CAL3	1	Water	NA	NA	20.1	10.03	01/30/2020	17:33
4	ICV	1	Water	NA	NA	20.2	7.00	01/30/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/30/2020	17:36
6	L1325-02	1	Solid	20.03	20	21.5	7.92	01/30/2020	17:38
7	L1325-02DUP	1	Solid	20.04	20	21.6	7.93	01/30/2020	17:39
8	L1325-04	1	Solid	20.02	20	21.3	5.62	01/30/2020	17:40
9	L1325-06	1	Solid	20.03	20	20.6	7.60	01/30/2020	17:42
10	L1325-08	1	Solid	20.04	20	21.1	7.78	01/30/2020	17:43
11	L1325-10	1	Solid	20.02	20	21.0	5.46	01/30/2020	17:45
12	L1333-06	1	Solid	20.03	20	20.8	5.72	01/30/2020	17:46
13	CCV2	1	Water	NA	NA	20.1	12.02	01/30/2020	17:48

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1333 WorkList ID : 135563 Department : Wet-Chemistry Date : 01-30-2020 17:01:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2020	Solid	L1325-02	Corrosivity	Cool 4 deg C	PSEG01	I43	346	01/30/2020	9045D
01/30/2020	Solid	L1325-04	Corrosivity	Cool 4 deg C	PSEG01	I43	200029	01/30/2020	9045D
01/30/2020	Solid	L1325-06	Corrosivity	Cool 4 deg C	PSEG01	I43	VNJ-232	01/30/2020	9045D
01/30/2020	Solid	L1325-08	Corrosivity	Cool 4 deg C	PSEG01	I43	72-11976	01/30/2020	9045D
01/30/2020	Solid	L1325-10	Corrosivity	Cool 4 deg C	PSEG01	I43	200059	01/30/2020	9045D
01/29/2020	Solid	L1333-06	Corrosivity	Cool 4 deg C	PSEG01	I43	TP-1-2-COMP	01/29/2020	9045D

Date/Time 01-30-2020 17:00
 Received by: JP
 Relinquished by: JP

Date/Time 01-30-2020 18:00
 Received by: JP
 Relinquished by: JP