



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB107452

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/30/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.00	01/30/2020	17:32
3	CAL3	1	Water	NA	NA	20.1	10.03	01/30/2020	17:33
4	ICV	1	Water	NA	NA	20.2	7.00	01/30/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/30/2020	17:36
6	L1329-01	1	Water	NA	NA	13.4	5.02	01/30/2020	17:37
7	L1329-02	1	Water	NA	NA	13.4	5.46	01/30/2020	17:38
8	L1329-03	1	Water	NA	NA	13.5	5.78	01/30/2020	17:40
9	L1329-04	1	Water	NA	NA	13.0	5.63	01/30/2020	17:42
10	L1329-05	1	Water	NA	NA	13.6	5.34	01/30/2020	17:44
11	L1329-06	1	Water	NA	NA	14.1	7.42	01/30/2020	17:45
12	L1329-07	1	Water	NA	NA	13.9	5.54	01/30/2020	17:46
13	L1329-08	1	Water	NA	NA	14.3	5.79	01/30/2020	17:48
14	L1329-09	1	Water	NA	NA	14.6	5.90	01/30/2020	17:50
15	L1329-16	1	Water	NA	NA	17.3	7.51	01/30/2020	18:00
16	CCV2	1	Water	NA	NA	20.1	12.01	01/30/2020	18:02
17	L1329-17	1	Water	NA	NA	17.2	6.42	01/30/2020	18:05
18	L1332-01	1	Water	NA	NA	13.6	7.40	01/30/2020	18:06
19	L1332-01DUP	1	Water	NA	NA	13.7	7.41	01/30/2020	18:07
20	CCV3	1	Water	NA	NA	20.1	2.01	01/30/2020	18:10

WORKLIST(Hardcopy Internal Chain)

18107462

WorkList Name : PH L1332

WorkList ID : 135560

Department : Wet-Chemistry

Date : 01-30-2020 15:36:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2020	Water	L1329-01	pH	Cool 4 deg C	PSEG01	I43	40364	01/30/2020	9040C
01/30/2020	Water	L1329-02	pH	Cool 4 deg C	PSEG01	I43	40366	01/30/2020	9040C
01/30/2020	Water	L1329-03	pH	Cool 4 deg C	PSEG01	I43	40673	01/30/2020	9040C
01/30/2020	Water	L1329-04	pH	Cool 4 deg C	PSEG01	I43	40633	01/30/2020	9040C
01/30/2020	Water	L1329-05	pH	Cool 4 deg C	PSEG01	I43	40631	01/30/2020	9040C
01/30/2020	Water	L1329-06	pH	Cool 4 deg C	PSEG01	I43	40632	01/30/2020	9040C
01/30/2020	Water	L1329-07	pH	Cool 4 deg C	PSEG01	I43	40672	01/30/2020	9040C
01/30/2020	Water	L1329-08	pH	Cool 4 deg C	PSEG01	I43	40630	01/30/2020	9040C
01/30/2020	Water	L1329-09	pH	Cool 4 deg C	PSEG01	I43	30564	01/30/2020	9040C
01/30/2020	Water	L1329-16	pH	Cool 4 deg C	PSEG01	I43	40230	01/30/2020	9040C
01/30/2020	Water	L1329-17	pH	Cool 4 deg C	PSEG01	I43	40098	01/30/2020	9040C
01/29/2020	Water	L1332-01	pH	Cool 4 deg C	HOME01	J43	KIT-SINK-1ST-DRAW	01/29/2020	9040C

Date/Time 01-30-2020 15:14:00
 Received by: JS
 Relinquished by: JS

Date/Time 01-30-2020 18:00
 Received by: JS
 Relinquished by: JS