



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107460

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/30/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	01/30/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/30/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	01/30/2020	17:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/30/2020	17:36
6	L1320-01	1	Solid	20.02	20	21.4	8.48	01/30/2020	17:40
7	L1320-01DUP	1	Solid	20.03	20	21.3	8.49	01/30/2020	17:41
8	L1329-10	1	Solid	20.04	20	19.4	6.76	01/30/2020	17:43
9	L1329-11	1	Solid	20.02	20	19.6	6.92	01/30/2020	17:45
10	L1329-12	1	Solid	20.03	20	19.5	7.86	01/30/2020	17:46
11	L1329-13	1	Solid	20.05	20	18.6	7.44	01/30/2020	17:49
12	L1329-14	1	Solid	20.02	20	19.9	7.52	01/30/2020	17:50
13	L1329-20	1	Solid	20.03	20	21.0	6.22	01/30/2020	18:00
14	CCV2	1	Water	NA	NA	20.1	12.01	01/30/2020	18:05

WORKLIST(Hardcopy Internal Chain)

107460

WorkList Name : PH S L1333

WorkList ID : 135562

Department : Wet-Chemistry

Date : 01-30-2020 17:00:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/30/2020	Solid	L1320-01	pH	Cool 4 deg C	PSEG01	I43	ETGI-367	01/30/2020	9045D
01/30/2020	Solid	L1329-10	pH	Cool 4 deg C	PSEG01	I43	40629	01/30/2020	9045D
01/30/2020	Solid	L1329-11	pH	Cool 4 deg C	PSEG01	I43	40628	01/30/2020	9045D
01/30/2020	Solid	L1329-12	pH	Cool 4 deg C	PSEG01	I43	40637	01/30/2020	9045D
01/30/2020	Solid	L1329-13	pH	Cool 4 deg C	PSEG01	I43	40636	01/30/2020	9045D
01/30/2020	Solid	L1329-14	pH	Cool 4 deg C	PSEG01	I43	40634	01/30/2020	9045D
01/30/2020	Solid	L1329-20	pH	Cool 4 deg C	PSEG01	I43	40510	01/30/2020	9045D

Date/Time 01-30-2020 17:10

Received by: JG

Relinquished by: DK

Date/Time 01-30-2020 18:25

Received by: DK

Relinquished by: JG