



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107462

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/31/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	01/31/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/31/2020	11:19
4	ICV	1	Water	NA	NA	20.1	7.01	01/31/2020	11:20
5	CCV1	1	Water	NA	NA	20.2	2.01	01/31/2020	11:22
6	L1334-01	1	Solid	20.03	20	21.6	13.50	01/31/2020	11:25
7	L1334-01DUP	1	Solid	20.04	20	21.7	13.51	01/31/2020	11:26
8	CCV2	1	Water	NA	NA	20.1	12.01	01/31/2020	11:30

WORKLIST(Hardcopy Internal Chain)

1810-462

Date : 01-31-2020 11:05:07

Department : Wet-Chemistry

WorkList ID : 135605

WorkList Name : CORROSIVITY L1334 S

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2020	Solid	L1334-01	✓ Corrosivity	Cool 4 deg C	ROYF02	J11	P001-WS004-01	01/29/2020	9045D

Date/Time 01-31-2020 11:10
 Received by: Jg
 Relinquished by: AM

Date/Time 01-31-2020 12:30
 Received by: AM
 Relinquished by: Jg