



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107463

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/31/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	01/31/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/31/2020	11:19
4	ICV	1	Water	NA	NA	20.1	7.01	01/31/2020	11:20
5	CCV1	1	Water	NA	NA	20.1	2.02	01/31/2020	11:35
6	L1334-06	1	Water	NA	NA	21.5	13.67	01/31/2020	11:36
7	L1334-07	1	Water	NA	NA	21.7	1.64	01/31/2020	11:38
8	L1334-08	1	Water	NA	NA	21.3	1.51	01/31/2020	11:39
9	L1334-08DUP	1	Water	NA	NA	21.4	1.52	01/31/2020	11:40
10	CCV2	1	Water	NA	NA	20.2	12.01	01/31/2020	11:44

WORKLIST(Hardcopy Internal Chain)

28107463

Date: 01-31-2020 11:05:54

Department: Wet-Chemistry

WorkList ID: 135606

WorkList Name: CORROSIONITY L1334 W

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2020	Water	L1334-06	✓ Corrosivity	Cool 4 deg C	ROYF02	J11	P001-WL002-01	01/29/2020	9040C
01/29/2020	Water	L1334-07	✓ Corrosivity	Cool 4 deg C	ROYF02	J11	P001-WL003-01	01/29/2020	9040C
01/29/2020	Water	L1334-08	✓ Corrosivity	Cool 4 deg C	ROYF02	J11	P001-WL005-01	01/29/2020	9040C

Date/Time 01-31-2020 11:10
 Received by: So
 Relinquished by: AM

Date/Time 01-31-2020 12:35
 Received by: AM
 Relinquished by: Jo