

7471B


UB/07464

Instrument ID = CV1

Linear

 μ Abs.:

61431



A= 0.0000e+000

B= 1.6302e-004

C= -3.3165e-002

Rho= 0.9999914

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	99	0.000	99				
0.2	0.200	0.211	0.011	1500	0.0 %	1500				
2.5	2.500	2.491	-0.009	15483	0.0 %	15483				
5.0	5.000	5.018	0.018	30985	0.0 %	30985				
7.5	7.500	7.515	0.015	46302	0.0 %	46302				
10.0	10.000	9.982	-0.018	61431	0.0 %	61431				

LB107464 INSTRUMENT ID:CV1

Method: 7471B		Operator: Admin		Date of Analysis: 31 Jan 2020 11:26:35				Date	Type	Type
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units				
0.0 - 1	50.0	99	-	0.0000	7471B	PPB	31 Jan 2020	11:28:45	S	Std
0.2 - 1	50.2	1500	-	0.2000	7471B	PPB	31 Jan 2020	11:31:02	S	Std
2.5 - 1	52.5	15483	-	2.5000	7471B	PPB	31 Jan 2020	11:33:19	S	Std
5.0 - 1	55.0	30985	-	5.0000	7471B	PPB	31 Jan 2020	11:35:36	S	Std
7.5 - 1	57.5	46302	-	7.5000	7471B	PPB	31 Jan 2020	11:37:52	S	Std
10.0 - 1	50.0	61431	-	10.0000	7471B	PPB	31 Jan 2020	11:40:08	S	Std
ICV95 - 1	ICV95	24340	3.9348	-	7471B	PPB	31 Jan 2020	11:44:22	U	SMPL
ICB95 - 1	ICB95	286	0.0135	-	7471B	PPB	31 Jan 2020	11:46:37	U	SMPL
CCV52 - 1	CCV52	31001	5.0207	-	7471B	PPB	31 Jan 2020	11:48:52	U	SMPL
CCB52 - 1	CCB52	289	0.0139	-	7471B	PPB	31 Jan 2020	11:51:07	U	SMPL
CRA - 1	CRA	1381	0.1920	-	7471B	PPB	31 Jan 2020	11:53:22	U	SMPL
HighStd - 1	HighStd	61009	9.9127	-	7471B	PPB	31 Jan 2020	11:55:36	U	SMPL
ChkStd - 1	ChkStd	42670	6.9230	-	7471B	PPB	31 Jan 2020	11:59:21	U	SMPL
PB126449BL - 1	PBS	22	-0.0296	-	7471B	PPB	31 Jan 2020	12:01:37	U	SMPL
PB126449BS - 1	LCSS	24527	3.9653	-	7471B	PPB	31 Jan 2020	12:03:53	U	SMPL
L1013-07 - 1	PL-05-012820	1100	0.1462	-	7471B	PPB	31 Jan 2020	12:06:09	U	SMPL
L1013-07DUP - 1	PL-05-012820DUP	1075	0.1421	-	7471B	PPB	31 Jan 2020	12:08:25	U	SMPL
L1013-07MS - 1	PL-05-012820MS	23541	3.8046	-	7471B	PPB	31 Jan 2020	12:20:42	U	SMPL
L1013-07MSD - 1	PL-05-012820MSD	24186	3.9097	-	7471B	PPB	31 Jan 2020	12:27:22	U	SMPL
L1278-01 - 1	TONNELE-AVE-01	29221	4.7305	-	7471B	PPB	31 Jan 2020	12:29:37	U	SMPL
CCV53 - 1	CCV53	32301	5.2326	-	7471B	PPB	31 Jan 2020	12:43:47	U	SMPL
CCB53 - 1	CCB53	216	0.0020	-	7471B	PPB	31 Jan 2020	12:46:01	U	SMPL
L1287-01 - 1	AOC-3-EXC-PIT-(0-5)-A	50209	8.1521	-	7471B	PPB	31 Jan 2020	12:48:16	U	SMPL
L1287-03 - 1	AOC-3-EXC-PIT-(0-5)-B	43944	7.1307	-	7471B	PPB	31 Jan 2020	12:50:31	U	SMPL
L1289-05 - 1	TP-26-27-COMP	3895	0.6018	-	7471B	PPB	31 Jan 2020	12:52:47	U	SMPL
L1289-13 - 1	TP-29-30-COMP	1190	0.1608	-	7471B	PPB	31 Jan 2020	12:55:02	U	SMPL
L1289-21 - 1	TP-32-33-COMP	1813	0.2624	-	7471B	PPB	31 Jan 2020	12:57:18	U	SMPL
L1289-29 - 1	TP-35-34-COMP	1308	0.1801	-	7471B	PPB	31 Jan 2020	12:59:35	U	SMPL
L1289-37 - 1	TP-3-COMP	1858	0.2697	-	7471B	PPB	31 Jan 2020	13:01:50	U	SMPL
PB126482BL - 1	PBS	166	-0.0061	-	7471B	PPB	31 Jan 2020	13:04:07	U	SMPL
PB126482BS - 1	LCSS	24737	3.9995	-	7471B	PPB	31 Jan 2020	13:06:24	U	SMPL
L1004-16 - 1	BU-01-012920	3580	0.5505	-	7471B	PPB	31 Jan 2020	13:08:39	U	SMPL
CCV54 - 1	CCV54	27942	4.5220	-	7471B	PPB	31 Jan 2020	13:10:55	U	SMPL
CCB54 - 1	CCB54	151	-0.0085	-	7471B	PPB	31 Jan 2020	13:13:10	U	SMPL
L1004-16DUP - 1	BU-01-012920DUP	3766	0.5808	-	7471B	PPB	31 Jan 2020	13:15:27	U	SMPL
L1004-16MS - 1	BU-01-012920MS	25705	4.1573	-	7471B	PPB	31 Jan 2020	13:17:42	U	SMPL
L1004-16MSD - 1	BU-01-012920MSD	25460	4.1174	-	7471B	PPB	31 Jan 2020	13:19:57	U	SMPL
L1014-11 - 1	OR-03-012920	5564	0.8739	-	7471B	PPB	31 Jan 2020	13:22:13	U	SMPL
L1296-01 - 1	RBR-51	4882	0.7627	-	7471B	PPB	31 Jan 2020	13:24:28	U	SMPL
L1298-01 - 1	RBR-54	1657	0.2370	-	7471B	PPB	31 Jan 2020	13:26:45	U	SMPL
L1307-01 - 1	SP-NORTH-SIDE	101599	16.5298	-	7471B	PPB	31 Jan 2020	13:29:01	U	SMPL
L1307-03 - 1	SP-SOUTH-SIDE	135774	22.1011	-	7471B	PPB	31 Jan 2020	13:31:18	U	SMPL
L1310-02 - 1	CTY-SB-0102-0-80	5954	0.9375	-	7471B	PPB	31 Jan 2020	13:33:37	U	SMPL
L1310-03 - 1	CTY-SB-10102-0-80	8143	1.2943	-	7471B	PPB	31 Jan 2020	13:35:55	U	SMPL
CCV55 - 1	CCV55	28171	4.5594	-	7471B	PPB	31 Jan 2020	13:38:11	U	SMPL
CCB55 - 1	CCB55	219	0.0025	-	7471B	PPB	31 Jan 2020	13:40:26	U	SMPL
L1313-01 - 1	RBR-36	1487	0.2092	-	7471B	PPB	31 Jan 2020	13:42:42	U	SMPL
L1313-02 - 1	124018-A	4011	0.6207	-	7471B	PPB	31 Jan 2020	13:44:57	U	SMPL
L1313-03 - 1	RBR-56	106	-0.0159	-	7471B	PPB	31 Jan 2020	13:47:12	U	SMPL
L1013-07LX5 - 1		225	0.0035	-	7471B	PPB	31 Jan 2020	13:49:28	U	SMPL
L1004-16LX5 - 1		749	0.0889	-	7471B	PPB	31 Jan 2020	13:51:43	U	SMPL
L1307-01DLX10 - 1		10534	1.6841	-	7471B	PPB	31 Jan 2020	13:54:00	U	SMPL
L1307-03DLX10 - 1		14120	2.2687	-	7471B	PPB	31 Jan 2020	13:56:16	U	SMPL
CCV56 - 1	CCV56	32228	5.2207	-	7471B	PPB	31 Jan 2020	14:04:45	U	SMPL
CCB56 - 1	CCB56	361	0.0257	-	7471B	PPB	31 Jan 2020	14:07:00	U	SMPL
PB126513BL - 1	PBS	-61	-0.0431	-	7471B	PPB	31 Jan 2020	15:03:40	U	SMPL
PB126513BS - 1	LCSS	20788	3.3558	-	7471B	PPB	31 Jan 2020	15:09:13	U	SMPL
L1318-01 - 1	RAINEY-PARK-BH-03-A	12133	1.9448	-	7471B	PPB	31 Jan 2020	15:11:28	U	SMPL
L1318-02 - 1	RAINEY-PARK-BH-03-B	28759	4.6552	-	7471B	PPB	31 Jan 2020	15:13:44	U	SMPL
L1318-02DUP - 1	RAINEY-PARK-BH-03-BDUP	28472	4.6084	-	7471B	PPB	31 Jan 2020	15:16:00	U	SMPL
L1318-02MS - 1	RAINEY-PARK-BH-03-BMS	53597	8.7044	-	7471B	PPB	31 Jan 2020	15:25:42	U	SMPL
L1318-02MSD - 1	RAINEY-PARK-BH-03-BMSD	51195	8.3128	-	7471B	PPB	31 Jan 2020	15:27:57	U	SMPL
CCV57 - 1	CCV57	33704	5.4614	-	7471B	PPB	31 Jan 2020	15:30:13	U	SMPL
CCB57 - 1	CCB57	83	-0.0196	-	7471B	PPB	31 Jan 2020	15:32:30	U	SMPL
L1318-03 - 1	RAINEY-PARK-BH-03-C	6834	1.0809	-	7471B	PPB	31 Jan 2020	15:34:46	U	SMPL
L1318-07 - 1	RAINEY-PARK-BH-04-A	1610	0.2293	-	7471B	PPB	31 Jan 2020	15:37:03	U	SMPL
L1318-08 - 1	RAINEY-PARK-BH-04-B	11471	1.8369	-	7471B	PPB	31 Jan 2020	15:39:20	U	SMPL
L1318-09 - 1	RAINEY-PARK-BH-04-C	2244	0.3327	-	7471B	PPB	31 Jan 2020	15:41:36	U	SMPL
L1318-13 - 1	RAINEY-PARK-BH-05-A	7082	1.1214	-	7471B	PPB	31 Jan 2020	15:43:52	U	SMPL
L1318-14 - 1	RAINEY-PARK-BH-05-B	4373	0.6797	-	7471B	PPB	31 Jan 2020	15:46:08	U	SMPL
L1318-15 - 1	RAINEY-PARK-BH-05-C	8574	1.3646	-	7471B	PPB	31 Jan 2020	15:48:23	U	SMPL
L1319-01 - 1	RAINEY-PARK-BH-06-C	3859	0.5959	-	7471B	PPB	31 Jan 2020	15:50:40	U	SMPL
L1319-05 - 1	RAINEY-PARK-BH-01-A	14394	2.3134	-	7471B	PPB	31 Jan 2020	15:52:56	U	SMPL
L1319-06 - 1	RAINEY-PARK-BH-01-B	16887	2.7198	-	7471B	PPB	31 Jan 2020	15:55:12	U	SMPL
CCV58 - 1	CCV58	33033	5.3520	-	7471B	PPB	31 Jan 2020	16:01:59	U	SMPL
CCB58 - 1	CCB58	163	-0.0066	-	7471B	PPB	31 Jan 2020	16:04:15	U	SMPL
L1319-07 - 1	RAINEY-PARK-BH-01-C	5089	0.7965	-	7471B	PPB	31 Jan 2020	16:06:31	U	SMPL
L1319-11 - 1	RAINEY-PARK-BH-02-A	2792	0.4220	-	7471B	PPB	31 Jan 2020	16:08:48	U	SMPL
L1319-12 - 1	RAINEY-PARK-BH-02-B	788	0.0953	-	7471B	PPB	31 Jan 2020	16:11:05	U	SMPL
L1319-13 - 1	RAINEY-PARK-BH-02-C	455	0.0410	-	7471B	PPB	31 Jan 2020	16:13:21	U	SMPL
L1319-17 - 1	RAINEY-PARK-BH-06-A	21582	3.4852	-	7471B	PPB	31 Jan 2020	16:15:37	U	SMPL
L1319-18 - 1	RAINEY-PARK-BH-06-B	14634	2.3525	-	7471B	PPB	31 Jan 2020	16:17:53	U	SMPL
L1330-02 - 1	TSP-1B	496	0.0477	-	7471B	PPB	31 Jan 2020	16:20:09	U	SMPL
PB126514BL - 1	PBS	11	-0.0314	-	7471B	PPB	31 Jan 2020	16:22:25	U	SMPL
PB126514BS - 1	LCSS	23827	3.8512	-	7471B	PPB	31 Jan 2020	16:24:41	U	SMPL
L1320-01 - 1	ETGI-367	1512	0.2133	-	7471B	PPB	31 Jan 2020	16:26:57	U	SMPL
CCV59 - 1	CCV59	29368	4.7545	-	7471B	PPB	31 Jan 2020	16:32:19	U	SMPL
CCB59 - 1	CCB59	239	0.0058	-	7471B	PPB	31 Jan 2020	16:34:34	U	SMPL
L1325-01 - 1	346	3729	0.5747	-	7471B	PPB	31 Jan 2020	16:36:50	U	SMPL
L1325-01DUP - 1	346DUP	3798	0.5860	-	7471B	PPB	31 Jan 2020	16:39:07	U	SMPL
L1325-01MS - 1	346MS	25401	4.1078	-	7471B	PPB	31 Jan 2020	16:41:24	U	SMPL
L1325-01MSD - 1	346MSD	25000	4.0424	-	7471B	PPB	31 Jan 2020	16:43:41	U	SMPL
L1325-03 - 1	200029	2522	0.3780	-	7471B	PPB	31 Jan 2020	16:45:57	U	SMPL

LB107464 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 Jan 2020 11:26:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
L1325-05 - 1	VNJ-232	3547	0.5451		-7471B	PPB	31 Jan 2020 16:48:12	U	SMPL
L1325-07 - 1	72-11976	4211	0.6533		-7471B	PPB	31 Jan 2020 16:50:28	U	SMPL
L1325-09 - 1	200059	2144	0.3164		-7471B	PPB	31 Jan 2020 16:52:44	U	SMPL
L1329-21 - 1	636-637	521	0.0518		-7471B	PPB	31 Jan 2020 16:55:00	U	SMPL
L1329-22 - 1	628-629	470	0.0435		-7471B	PPB	31 Jan 2020 16:57:16	U	SMPL
CCV60 - 1	CCV60	32864	5.3244		-7471B	PPB	31 Jan 2020 16:59:32	U	SMPL
CCB60 - 1	CCB60	395	0.0312		-7471B	PPB	31 Jan 2020 17:01:47	U	SMPL
L1331-01 - 1	C-1_013020	2840	0.4298		-7471B	PPB	31 Jan 2020 17:04:04	U	SMPL
L1331-02 - 1	C-2_013020	3501	0.5376		-7471B	PPB	31 Jan 2020 17:06:21	U	SMPL
L1331-03 - 1	C-3_013020	883	0.1108		-7471B	PPB	31 Jan 2020 17:08:37	U	SMPL
L1333-05 - 1	TP-1-2-COMP	1217	0.1652		-7471B	PPB	31 Jan 2020 17:10:53	U	SMPL
L1318-02LX5 - 1		5656	0.8889		-7471B	PPB	31 Jan 2020 17:13:10	U	SMPL
L1325-01LX5 - 1		603	0.0651		-7471B	PPB	31 Jan 2020 17:15:26	U	SMPL
CCV61 - 1	CCV61	30628	4.9599		-7471B	PPB	31 Jan 2020 17:17:42	U	SMPL
CCB61 - 1	CCB61	-96	-0.0488		-7471B	PPB	31 Jan 2020 17:19:59	U	SMPL