



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107465

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP81698
Turbidity Calibration std, 0NTU	WP81697
Turbidity Calibration std, 20NTU	WP81693
Turbidity Calibration std, 40NTU	WP81692
Turbidity Calibration std, 1NTU	WP81696
Turbidity Calibration std, 10NTU	WP81694
Turbidity Calibration std, 5NTU	WP81695
10 NTU Standard 500 ml	W2489

Intercept: 0.0712

Slope: 0.8862

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.169	01/31/2020	10:30
2	CAL2	1	1	1.090	01/31/2020	10:32
3	CAL3	5	1	4.560	01/31/2020	10:34
4	CAL4	10	1	8.960	01/31/2020	10:36
5	CAL5	20	1	17.200	01/31/2020	10:38
6	CAL6	40	1	35.800	01/31/2020	10:40



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107465

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.350	10.470	01/31/2020	10:43
2	ICB		1	0.159	0.099	01/31/2020	10:46
3	CCV1	10	1	9.060	10.143	01/31/2020	10:49
4	CCB1		1	0.150	0.089	01/31/2020	10:55
5	LB107465BL		1	0.160	0.100	01/31/2020	10:58
6	L1161-12		1	0.629	0.629	01/31/2020	11:02
7	L1332-01		1	0.155	0.095	01/31/2020	11:05
8	L1332-01DUP		1	0.160	0.100	01/31/2020	11:11
9	CCV2	10	1	8.980	10.053	01/31/2020	11:14
10	CCB2		1	0.159	0.099	01/31/2020	11:17

LB107465

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity013120 WorkList ID : 135595 Department : Wet-Chemistry Date : 01-31-2020 09:15:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/18/2020	Water	L1161-12	Turbidity	Cool 4 deg C	CHEM02	QA Office	MDL-WATER-06-QT1-2020	01/16/2020	SM2130 B
01/31/2020	Water	L1332-01	Turbidity	Cool 4 deg C	HOME01	J43	KIT-SINK-1ST-DRAW	01/29/2020	SM2130 B

Date/Time 01-31-20 10:15
 Received by: AK
 Relinquished by: JP

Date/Time 01-31-20 11:30
 Received by: CP
 Relinquished by: AK