

LB107477

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

Instrument ID : Konelab

1/31/2020 15:14

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	99.340	0.0	0.106	
ICB1	-0.168	0.0	0.003	
CCV1	248.658	0.0	0.261	
CCB1	-0.457	0.0	0.002	
PB126437BL	-0.167	0.0	0.003	
PB126437BS	100.608	0.0	0.107	
MIDPB126437	256.661	0.0	0.270	
L1161-12	2.073	0.0	0.005	
CCV2	253.539	0.0	0.266	
CCB2	-0.281	0.0	0.003	

N	10
Mean	95.981
SD	115.4027
CV%	120.24

AK

Aquakem v. 7.2AQ1

Results from time period:

Fri Jan 31 14:45:56 2020

Fri Jan 31 15:04:57 2020

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.8464	µg/l	1/31/2020 9:23:56	
5.OPPBCN	A	Total CN	P	4.1422	µg/l	1/31/2020 9:23:57	
10PPBCN	A	Total CN	P	9.2035	µg/l	1/31/2020 9:23:58	
50PPBCN	A	Total CN	P	50.3314	µg/l	1/31/2020 9:23:59	
100PPBCN	A	Total CN	P	101.2301	µg/l	1/31/2020 9:24:00	
250PPBCN	A	Total CN	P	252.3876	µg/l	1/31/2020 9:24:01	
500PPBCN	A	Total CN	P	498.5515	µg/l	1/31/2020 9:24:02	
ICV1	S	Total CN	P	99.3397	µg/l	1/31/2020 14:45:56	
ICB1	S	Total CN	P	-0.1679	µg/l	1/31/2020 14:45:57	
CCV1	S	Total CN	P	248.6584	µg/l	1/31/2020 15:04:50	
CCB1	S	Total CN	P	-0.4566	µg/l	1/31/2020 15:04:51	
PB126437BL	S	Total CN	P	-0.1668	µg/l	1/31/2020 15:04:52	
PB126437BS	S	Total CN	P	100.6085	µg/l	1/31/2020 15:04:53	
MIDPB126437	S	Total CN	P	256.6614	µg/l	1/31/2020 15:04:54	
L1161-12	S	Total CN	P	2.0734	µg/l	1/31/2020 15:04:55	
CCV2	S	Total CN	P	253.5389	µg/l	1/31/2020 15:04:56	
CCB2	S	Total CN	P	-0.2809	µg/l	1/31/2020 15:04:57	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AD

1/31/2020 14:31

Instrument ID : Konelab


Test Total CN

Accepted 1/31/2020 9:25

Factor 962.5
Bias 0.003

Coeff. of det. 0.999944

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.002	-0.8464	0.0000	
2	5.0PPBCN	0.007	4.1422	5.0000	
3	10PPBCN	0.013	9.2035	10.0000	
4	50PPBCN	0.055	50.3314	50.0000	
5	100PPBCN	0.108	101.2301	100.0000	
6	250PPBCN	0.265	252.3876	250.0000	
7	500PPBCN	0.521	498.5515	500.0000	