

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB107480

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP81685
calibration std. phosphate 0.5 ppm	WP81684
calibration std. phosphate 0.3 ppm	WP81683
calibration std. phosphate 0.1 ppm	WP81682
calibration std. phosphate 0.05 ppm	WP81681
calibration std. 0 ppm	WP81680
phosphate ICV std.	WP81687
phosphate CCV std.	WP81686
5N sulfuric acid	WP79628
Combined reagent	WP81689
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP81711
Sodium hydroxide, 1N	WP77314

Intercept: -0.0008Slope: 0.6496Regression: 0.999856

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/31/2020	13:46
2	CAL2	0.05	1	50	50	0.033	01/31/2020	13:46
3	CAL3	0.10	1	50	50	0.066	01/31/2020	13:47
4	CAL4	0.30	1	50	50	0.186	01/31/2020	13:47
5	CAL5	0.50	1	50	50	0.328	01/31/2020	13:48
6	CAL6	1.00	1	50	50	0.649	01/31/2020	13:48



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB107480

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.304	0.469	01/31/2020	13:49
2	ICB		1	50	50	0.000	0.001	01/31/2020	13:49
3	CCV1	0.50	1	50	50	0.325	0.502	01/31/2020	13:50
4	CCB1		1	50	50	0.000	0.001	01/31/2020	13:50
5	PB126511BL		1	50	50	0.000	0.001	01/31/2020	13:51
6	PB126511BS	0.50	1	50	50	0.316	0.488	01/31/2020	13:52
7	L1161-07		1	50	50	0.013	0.021	01/31/2020	13:53
8	L1161-08		1	50	50	0.027	0.043	01/31/2020	13:53
9	L1272-01		1	50	50	0.405	0.625	01/31/2020	13:53
10	L1272-01DUP		1	50	50	0.403	0.622	01/31/2020	13:54
11	L1272-01MS	0.50	1	50	50	0.748	1.153	01/31/2020	13:55
12	L1272-01MSD	0.50	1	50	50	0.740	1.140	01/31/2020	13:55
13	L1272-05		1	50	50	0.806	1.242	01/31/2020	13:56
14	L1272-01MS	0.50	2	50	50	0.365	0.563	01/31/2020	14:19
15	CCV2	0.50	1	50	50	0.326	0.503	01/31/2020	14:21
16	CCB2		1	50	50	0.000	0.001	01/31/2020	14:21
17	L1272-01MSD	0.50	2	50	50	0.361	0.557	01/31/2020	14:22
18	L1272-05		2	50	50	0.458	0.706	01/31/2020	14:23
19	CCV3	0.50	1	50	50	0.327	0.505	01/31/2020	14:23
20	CCB3		1	50	50	0.000	0.001	01/31/2020	14:24